

Automatic Optic Nerve Measurement: A New Tool to Standardize Optic Nerve Assessment in
Ultrasound B-Mode Images

Original

Automatic Optic Nerve Measurement: A New Tool to Standardize Optic Nerve Assessment in Ultrasound B-Mode Images / Meiburger, K. M.; Naldi, A.; Michielli, N.; Coppo, L.; Fassbender, K.; Molinari, F.; Lochner, P.. - In: ULTRASOUND IN MEDICINE AND BIOLOGY. - ISSN 0301-5629. - (2020). [10.1016/j.ultrasmedbio.2020.01.034]

Availability:

This version is available at: 11583/2809626 since: 2020-04-08T20:38:12Z

Publisher:

Elsevier USA

Published

DOI:10.1016/j.ultrasmedbio.2020.01.034

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

Elsevier postprint/Author's Accepted Manuscript

© 2020. This manuscript version is made available under the CC-BY-NC-ND 4.0 license
<http://creativecommons.org/licenses/by-nc-nd/4.0/>. The final authenticated version is available online at:
<http://dx.doi.org/10.1016/j.ultrasmedbio.2020.01.034>

(Article begins on next page)

1 **AUTomatic Optic Nerve MeAsurement (AUTONoMA): a new tool to**
2 **standardize the optic nerve assessment in ultrasound B-mode images**

3

4 Kristen M. Meiburger^a, Andrea Naldi^b, Nicola Michielli^a, Lorenzo Coppo^c, Klaus
5 Fassbender^d, Filippo Molinari^a and Piergiorgio Lochner^d

6

7 **Author affiliations:**

8 ^aPoliTo^{BIO}Med Lab, Biolab, Department of Electronics and Telecommunications,
9 Politecnico di Torino, Turin, Italy

10 ^bDepartment of Neuroscience "Rita Levi Montalcini", University of Turin, Turin, Italy

11 ^cNeurology Unit, Department of Translational Medicine, University of Piemonte Orientale,
12 Novara, Italy

13 ^dDepartment of Neurology, Saarland University Medical Center, Homburg, Germany

14

15 **Corresponding author:**

16 Kristen M. Meiburger

17 Politecnico di Torino

18 Biolab, Department of Electronics and Telecommunications

19 Corso Duca degli Abruzzi, 24

20 10129 Torino, ITALY

21 Phone: +39 011 090 4207

22 E-mail: kristen.meiburger@polito.it

23

24

25 **Abstract**

26 Transorbital sonography provides reliable information about the estimation of intracranial
27 pressure by measuring the optic nerve sheath diameter (ONSD), while the optic nerve
28 (ON) diameter (OND) may reveal ON atrophy in multiple sclerosis patients. Here, an
29 AUTomatic Optic Nerve MeAsurement (AUTONoMA) system for OND and ONSD
30 assessment in ultrasound B-mode images based on deformable models is presented.
31 The automated measurements were compared to manual ones obtained by two
32 operators, with no significant differences. AUTONoMA correctly segmented the ON and
33 its sheath in 71 out of 75 images. The mean error compared with the expert operator was
34 0.06 ± 0.52 mm and 0.06 ± 0.35 mm for the ONSD and OND respectively. The agreement
35 between operators and AUTONoMA was good and a positive correlation between the
36 readers and the algorithm with errors comparable with the inter-operator variability was
37 found. The AUTONoMA system may allow a standardization of OND and ONSD
38 measurements, reducing manual evaluation variability.

39

40 **Keywords:** Ultrasound, Optic nerve segmentation, Intracranial pressure, Optic nerve
41 diameter, Optic nerve sheath diameter.

42

43

44 Introduction

45 Transorbital sonography (TOS) is a promising technique for the non-invasive
46 evaluation of the optic nerve (ON) structures. This tool is particularly versatile and can be
47 performed both in remote, prehospital setting and hospital context, either in invasive or
48 non-invasive departments (Houzé-Cerfon et al. 2018; Lochner et al. 2015). The main use
49 of TOS concerns the assessment of the optic nerve sheath diameter (ONSD) for the
50 estimation and monitoring of increased intracranial pressure (ICP), particularly when the
51 invasive referenced methods are contraindicated or unavailable (Goeres et al. 2016;
52 Robba et al. 2015; Soliman et al. 2018). Moreover, TOS can be useful to detect ON
53 atrophy in patients with multiple sclerosis (Carraro et al. 2014). **In the past, transorbital**
54 **sonography has been performed by using amplitude-mode (A) standardized**
55 **ultrasonography, which provides simple displays plotted as a series of peaks whose**
56 **height represents the depth of the echoing structure from the transducer (Ossoinig 1979;**
57 **Schroeder 1976). Due to software improvement and the development of higher frequency**
58 **probes, Brightness-mode (B) scan replaced the A-Mode sonography. The advantages of**
59 **B-Mode sonography includes the generation of a two-dimensional image, allowing a**
60 **better topography of the tissue with direct visualization of lesions.** The current application
61 fields of **B-Mode** TOS in the clinical practice have been recently described (Lochner et al.
62 2019).

63 A good intra and interobserver reproducibility using high-frequency (>7.5 MHz) linear
64 probe, which allows a lateral spatial resolution <0.4 mm, can be obtained for the
65 ultrasonographic assessment of optic nerve diameter (OND) and ONSD (Bäuerle et al.
66 2012; Lochner et al. 2014; Lochner et al. 2018a). Although this, the manual evaluation of

67 OND and ONSD can be affected by the operator's experience and artefactual images
68 (Ballantyne et al. 2002; Copetti and Cattarossi 2009). In addition, different methods are
69 currently described in literature for the ONSD evaluation, leading to possible
70 misunderstanding in the results interpretation (Bloria et al. 2019).

71 Even if a greater experience or a continuous training have demonstrated to reduce
72 operator variability, for a better use of the technique, a unique model of measurements
73 and a standardization of the method are required (Zeiler et al. 2013; Zeiler et al. 2014).

74 The development of computerized automated systems for the segmentation of structures
75 in B-mode ultrasound images is an auspicious research field that may help reduce
76 thereupon the operator-dependency, accelerate the acquisition time and mitigate the
77 issue of inter-operator variability (Meiburger et al. 2018). In this context, Gerber et al.
78 (Gerber et al. 2017) developed an algorithm to automatically estimate the ONSD from 23
79 ocular ultrasound images and on an eye phantom using 3D-printed optic nerves
80 embedded under gelatin orbs, and Soroushmehr et al. (Soroushmehr et al. 2019)
81 developed a method based on super-pixel analysis to measure the ONSD in 50
82 ultrasound images. However, to the best of our knowledge, except from these works
83 which employed a smaller dataset of in-vivo images and estimated only the ONSD, there
84 are no described methods focused on a completely automatic segmentation of the optic
85 nerve and optic nerve sheath in ultrasound B-mode images in a series of patients affected
86 by neurological diseases with increased ICP and healthy subjects.

87 Therefore, the aim of this work is to present and validate a completely automatic system
88 for measuring the OND and ONSD, requiring no interaction with the user.

89

90 **Materials and Methods**

91 The measurement of the OND and ONSD with TOS is based on the difference in
92 echogenicity and morphology of the different retro-orbital structures. The developed
93 algorithm is based on the assumption that the ultrasound image presents hypoechoic
94 structures like the vitreous, the inside of the optic nerve and the arachnoid, and
95 hyperechoic structures like pia and dura mater and the surrounding adipose tissue. The
96 anterior part of the optic nerve is depicted in an axial plane showing the papilla and the
97 optic nerve in its longitudinal course. ONSD and OND are assessed 3 mm behind the
98 papilla (Helmke and Hansen 1996) and should be calculated perpendicularly to the optic
99 nerve centerline.

100 The OND is typically measured manually as the distance between the right profile
101 of the optic nerve and the left one. To measure the ONSD, we quantified the distance
102 between the external borders of the hyperechogenic area surrounding the optic nerve
103 (Ertl et al. 2014), as shown in Fig. 1.

104

105 ***Image acquisition and database***

106 A total of 75 images were included in this study: 30 images came from 15 patients
107 who were diagnosed either with primary or secondary intracranial hypertension (IH)
108 according to the current diagnostic criteria (Friedman et al. 2013) and 45 images from 23
109 healthy controls. The study was approved by the local ethical committee (Bolzano,
110 20/2014) and all participants provided written informed consent before being included. All
111 images were acquired by an expert neurosonologist with more than 10 years of

112 experience in TOS using a Vivid 7 sonography system with a 7 to 11 MHz linear array
113 probe with a central frequency of 10 MHz (GE Healthcare, Milwaukee, Wisconsin).

114 For image acquisition, a standard protocol was followed. Specifically, the patient
115 was asked to lie in a supine position on a bed with the head reclined at a 20°-30° angle.
116 With the patient's eyes closed, the linear array ultrasound probe was gently placed on the
117 closed eyelids (never in direct contact with the cornea or sclera), and the image was
118 acquired. All images were exported from the ultrasound device and transferred to a
119 workstation for offline processing.

120

121

122 ***AUTONoMA architecture***

123 An overview of our proposed AUTomatic Optic Nerve MeAsurement (AUTONoMA)
124 system is presented in Fig. 2. It consists of a computer aided diagnosis (CAD) system
125 that takes a B-mode image obtained from transorbital ultrasonography and gives forth an
126 automated measurement of the optic nerve diameter and the optic nerve sheath diameter,
127 without requiring any interaction from the user. The proposed system can be summarized
128 in two main automatic steps:

129

- 130 1. Stage I: coarse localization of the **region-of-interest** through the automatic
131 recognition of the ocular bulb profile and the optic nerve centerline tracing.
- 132 2. Stage II: fine segmentation of the optic nerve and the optic nerve sheath through
133 dual snakes and automatic measurement of OND and ONSD.

134

135 Stage I: coarse localization of ocular bulb and optic nerve centerline

136 In order to accurately locate the ocular bulb within the ultrasound image frame, a
137 preprocessing step is first necessary to isolate the ultrasound information from the entire
138 image frame. To do so, an automatic image cropping step was developed, using
139 morphological operations and gradients. Fig. 3a shows the original image and Fig. 3b
140 shows the automatically cropped image.

141 The automatic recognition of the ocular bulb is then done on the cropped image.
142 First of all, the image is sharpened by summing the original image with the image obtained
143 with the First Order Absolute Moment (FOAM), an edge operator which has been applied
144 previously in ultrasound images (Faita et al. 2006). Subsequently, a gaussian derivative
145 filter (sigma = 7) was applied to the image (Fig. 3c). In the obtained image, the bottom
146 border of the optical bulb is automatically located through a column-wise heuristic search.
147 The beginning of the column-wise search region is found by locating the first pixel (starting
148 from the top of the image) that is above a certain intensity and that presents a reasonably
149 large hypoechogenic region above it (i.e., the bulb) (Fig. 3c). The point of the optical bulb
150 boundary for the analyzed column is then automatically located by finding the first
151 discontinuity from a hypoechoic zone to a hyperechoic zone on the B-mode cropped
152 image. The profile obtained by analyzing each column, $AUTONoMA_{bulb}$, is then
153 interpolated to give forth the final segmentation of the ocular bulb (Fig. 3d).

154 The identification of the bottom boundary of the ocular bulb makes the search for
155 an estimation of the optic nerve centerline significantly easier by limiting the search of the
156 optic nerve within a specific area of the image. Specifically, a gaussian derivative filter
157 was again applied to the ultrasound image and the optic nerve centerline was located

158 thanks to a heuristic search below the found AUTONoMA_{bulb} profile similar to the one
159 previously described but considered row-wise. The obtained centerline, AUTONoMA_{ONc},
160 is shown in Fig. 3e.

161

162 Stage II: fine segmentation of optic nerve and optic nerve sheath

163 The fine segmentation of the optic nerve and the optic nerve sheath is done by
164 implementing a dual snake model, similar to the one presented by Molinari et al. (Molinari
165 et al. 2012b; Molinari et al. 2012a). Since both the optic nerve diameter and the optic
166 nerve sheath diameter are of clinical interest, two different dual snake models were
167 developed: one for the calculation of the optic nerve diameter, and the other for the optic
168 nerve sheath diameter.

169 As all active contour models, the dual snake algorithm requires a first initialization
170 of the snakes, which then evolve in time and adapt to the optic nerve and sheath
171 boundaries. The snakes initialization and evolution are described in the following
172 paragraphs.

173

174 *Snakes initialization*

175 The snakes initialization can be summarized as follows: 1) starting from the located
176 optic nerve centerline, the ON dual snake (ONDS) model was initialized by locating the
177 rough nerve boundary; 2) similarly, the optic nerve sheath dual snake (ONSDS) model
178 was initialized by locating the rough sheath boundary starting from the rough nerve
179 boundary located in the previous step 1.

180 The rough boundaries of both the optic nerve and the optic nerve sheath, hence
181 the snakes initialization, were located thanks to a row-wise heuristic search on the original
182 image filtered with a gaussian derivative filter in two directions. Briefly, starting from the
183 centerline/optic nerve boundary going outwards, the first pixel in the gaussian derivative
184 filtered image that is higher than a specific threshold is taken as the candidate point for
185 the optic nerve/optic nerve sheath. The snake initialization then is taken as joining
186 together all candidate points.

187 Fig. 4a and 4b shows the snakes initialization for the ONDS and ONSDS,
188 respectively.

189

190 *Snakes evolution*

191 Once the snakes are initialized, the dual snake models ($v(s)$) then evolve in time
192 thanks to three energy models: the internal, external, and mutual interaction energies.
193 The internal energy serves to constrain the shape of the contour and prevents the active
194 contour from presenting an excessive curvature, which is especially necessary in this
195 clinical application, in which the optic nerve and optic nerve sheath are represented by
196 more or less straight lines. This energy is defined as:

$$E_{int}(v(s)) = \int_0^1 \alpha |v'(s)| ds \quad (1)$$

197

198 where s is the curvilinear coordinate on the image, $v'(s)$ is the first-order derivative of the
199 snake curve $v(s)$ and α is a parameter used to give a specific weight to the internal
200 energy, controlling the curvature of the snake.

201 The external energy is what attracts the snake model toward the image
202 discontinuities. This energy is defined as:

$$E_{ext}(v(s)) = - \int_0^1 \beta e(v(s)) ds \quad (2)$$

203
204 where β is a parameter used to give a specific weight to the external energy and the
205 functional $e(x, y)$ is a first order gaussian derivative filter, an edge operator that has been
206 used in numerous ultrasound clinical applications (Caresio et al. 2017).

207 The mutual interaction energy, which can be considered as a second term of
208 external energy, is necessary to ensure that the two models of the dual snake do not
209 either collapse on one another or converge. So, this energy is inversely proportional to
210 the distance between the two curves (the left and right snake, $v_L(s)$ and $v_R(s)$,
211 respectively) and is defined as:

$$E_{mut}(v(s)) = \int_0^1 \gamma \frac{1}{|v_R(s) - v_L(s)|} ds \quad (3)$$

212
213 where γ is a parameter used to give a specific weight to the mutual energy.

214 The values of the parameters used for each of the dual snake models are shown
215 in Tab. 1. As can be seen, the external energy and mutual energy parameters are
216 dependent on the conversion factor (CF), expressed in $mm/pixel$, in order to make the
217 models independent of both zooming and of the ultrasound device used to acquire the
218 images. The value of CF_{base} was equal to $0.116 mm/pixel$.

219 The final segmentation of the optic nerve and the optic nerve sheath is shown in
220 Fig. 4c and 4d, respectively.

221

222 *Calculation of the OND and ONSD*

223 Once the optic nerve and optic nerve sheath are correctly segmented, the
224 diameters of the two structures (OND and ONSD) were automatically measured. This
225 was done by using the optic nerve centerline that was found automatically and locating
226 the point that is 3 mm behind the optic bulb. From here, the Centerline Distance (Saba et
227 al. 2012) between the two final snake models was calculated to give forth the final OND
228 and ONSD values (Fig. 4e). In order to reduce the variability of the final OND and ONSD
229 measurements, the centerline distance was calculated right at 3 mm behind the optic
230 bulb, slightly before 3 mm, and slightly after 3 mm, and the average distance was taken
231 to be the final diameter measurement.

232 The AUTONoMA system was developed in Matlab and showed an average
233 computational time of 2 seconds for processing a single image, providing an almost real-
234 time analysis.

235

236 ***Performance evaluation***

237 In order to validate the results of the developed AUTONoMA algorithm, different
238 performance evaluation metrics were used.

239 First of all, the OND and ONSD measurements that were obtained automatically
240 were compared with the manual measures of an expert with more than 10 years of
241 experience in transorbital ultrasonography and a non-expert operator (referenced as Op1
242 and Op2 from here on out, respectively), considered as ground truth. To do the manual
243 measurement, an in-house program in Matlab was developed to allow adequate zooming

244 of the image, and the subsequent manual tracing of the optic nerve centerline. Using the
 245 calibration factor, the perpendicular line at 3mm was drawn and the operator was asked
 246 to use the mouse to measure the OND and ONSD at the correct depth. So, for each
 247 image, the error between the automatic computer-based measure and the ground truth
 248 measure was calculated. Three types of error were used to describe the overall system
 249 performance: the mean error (defined as the mean difference between the manual
 250 measure and the automatic one), the mean absolute error (MAE) and the mean squared
 251 error (MSE), along with the respective standard deviations. Another parameter, the Figure
 252 of Merit (FoM), which characterizes the overall performances of the algorithm, was
 253 calculated. This parameter is defined as:

$$FoM_{OND} = 100 - \left| \frac{mean(OND_{auto}) - mean(OND_{man})}{mean(OND_{man})} \right| \cdot 100 \quad (4)$$

254

$$FoM_{ONSD} = 100 - \left| \frac{mean(ONSD_{auto}) - mean(ONSD_{man})}{mean(ONSD_{man})} \right| \cdot 100 \quad (5)$$

255

256 where $mean(OND_{auto})$ and $mean(ONSD_{auto})$ are respectively the average OND and
 257 ONSD values found automatically, and $mean(OND_{man})$ and $mean(ONSD_{man})$ are the
 258 average OND and ONSD values measured manually, respectively.

259 Moreover, we calculated the correlation coefficient and the 95% confidence interval
 260 between the ground truth diameter values and the automated diameter values. Finally, to
 261 determine if the automatic and manual measurements present a statistically significant
 262 difference between the measurements or not, the Wilcoxon signed rank test was used.

263 In order to assess inter-operator variability in OND and ONSD measurements, the
264 correlation between the manual measurements was also calculated. For each image, the
265 manual measurements were obtained offline and independently by the two operators
266 involved (both blinded with regard to AUTONoMA performance). The developed
267 AUTONoMA system is completely automated and independent from the user; therefore
268 the system does not present any measurement variability.

269

270 **Results**

271

272 The proposed AUTONoMA system was able to **process** 71 out of 75 images,
273 presenting a 95% success rate. Fig. 5 shows some example segmentation results
274 obtained with the AUTONoMA system, whereas Fig. 6 shows two examples of images
275 that were not able to be processed automatically.

276 The performance values for the optic nerve and the optic nerve sheath diameter
277 are reported in Tab. 2 and Tab. 3, respectively. No **statistically** significant differences were
278 observed between AUTONoMA and the operators for the OND values ($p > 0.05$), whereas
279 a **statistically** significant difference was found for the ONSD values between only
280 AUTONoMA and the inexperienced operator ($p < 0.05$). Considering the OND measurements,
281 the automatic algorithm gave forth mean errors equal to $0.06 \pm 0.35 \text{ mm}$ and $0.05 \pm$
282 0.38 mm when compared with Op1 and Op2, respectively. In both cases the algorithm
283 underestimated the measure. The FoM was equal to 98.2% when comparing results with
284 Op1 and equal to 98.3% when considering Op2. When considering the ONSD, on the

285 other hand, the mean error compared to Op1 and Op2 was found to be equal to $0.06 \pm$
286 0.52 mm and $-0.37 \pm 0.55 \text{ mm}$, and the FoM was 99.0% and 93.5%, respectively.

287 The Pearson correlation coefficient, the 95% confidence interval and the p-value
288 between the automatic measure, both for the OND and the ONSD, and the manual one
289 performed by Op1 and Op2 are reported in Tab. 4. A statistically significant correlation
290 between our developed AUTONoMA algorithm and the manual operators was found,
291 showing p-values ≤ 0.05 in all cases, considering both the OND and the ONSD. The inter-
292 operator variability also showed a statistically significant correlation (p-value ≤ 0.05 ,
293 considering both OND and ONSD).

294 The Bland-Altman plots of the AUTONoMA optic nerve diameter and optic nerve
295 sheath diameter compared to Op1 and Op2 are shown in Fig. 7a. The Bland-Altman plots
296 related to the inter-operator analysis for the OND and ONSD are reported in Fig. 7b. It
297 can be appreciated that there is an absence of any visible bias and the automated
298 measurements were all close to the manually measured values.

299

300

301 **Discussion**

302

303 Apart from a recent pilot study presented by Gerber et al. (Gerber et al. 2017), this
304 is the first work that proposes an automatic optic nerve system to calculate both the OND
305 and the ONSD. We used a three times larger dataset of ocular ultrasound images,
306 comparing the automated measurements with those of two investigators with different
307 expertise who independently examined both parameters.

308 The main findings of our work are as follows:

309 Firstly, the developed AUTONoMA algorithm is fully automated and was able to
310 process 95% of the images present in the dataset. On further analysis, it was found that
311 the automatic algorithm provided a segmentation that the manual expert deemed as
312 acceptable in all cases except for 5 images (7%), where the AUTONoMA segmentations
313 diverged from the actual ON and ONS borders.

314 Secondly, the mean value of ONSD obtained from AUTONoMA, $6.2 \pm 0.6 \text{ mm}$, is very
315 similar to the ONSD value achieved by the expert operator, $6.2 \pm 0.6 \text{ mm}$ ($p = 0.28$), and
316 significantly different from the inexperienced operator, $5.8 \pm 0.6 \text{ mm}$ ($p < 0.05$). Moreover, the
317 mean absolute and mean squared errors exclude a systematic error from AUTONoMA.
318 Similarly, the OND measurements obtained from AUTONoMA were not significantly
319 different from those achieved by both the operators ($p = 0.08$ and $p = 0.21$, respectively).
320 Thirdly, regarding the inter-observer reliability according to the Bland-Altman analysis, we
321 found a good agreement between the operators. Moreover, the difference of
322 measurements of ONSD is inferior to the intrinsic error of the machine (Ballantyne et al.
323 2002) and comparable with the inter-operator reproducibility reported in prior studies
324 (Bäuerle et al. 2012; Lochner et al. 2014).

325 Finally, AUTONoMA calculates the OND/ONSD value simultaneously and in a very
326 short time, approximately two seconds per image. Manual measurements take about 30
327 seconds for each image, hence the automatic algorithm provides a result fifteen times
328 faster than an expert operator. Moreover, since the developed system is completely
329 automatic and independent from the user, there is no OND/ONSD measurement
330 variability.

331 Translating into clinical practice, the AUTONoMA system may represent the first
332 step to reduce the wide variability of ONSD and OND measurements currently described
333 in literature. These differences reflect the operator's experience, the use of different
334 ultrasonographic machines, and a non-homogeneous and standardized method for
335 image acquisition and measurements (Bloria et al. 2019). The presence and use of an
336 automated system such as AUTONoMA could – at least in part – mitigate and minimize
337 these differences, promoting a more comparable interpretation of results among studies.

338 To date, both OND and ONSD are used to study neurological conditions that imply
339 variations in their value. Meanwhile, the spectrum of applications for TOS in the context
340 of neurological diseases has progressively extended (Lochner et al. 2019). This is due to
341 the versatility of the ultrasonographic evaluation because of its availability,
342 inexpensiveness, repeatability and bedside use. For these reasons, the sonographic
343 assessment of the ONSD is considered an alternative to the invasive evaluation for the
344 estimation of increased ICP, especially in pre-hospital settings or when radiological or
345 neurosurgical care are not available or contraindicated (Robba et al. 2018). However, a
346 clear cut-off value to identify intracranial hypertension is not available, probably due to
347 differences in sex, ethnicity, age, body mass index and technical limitations cited above;
348 also anatomical factors or previous ocular or cerebral pathologies may be implied in
349 generating a variability of optic nerve structures (Bäuerle et al. 2016; Naldi et al. 2019;
350 Wang et al. 2016). Thus, the emergent concept of monitoring ONSD values from a basal
351 level is taking place: if a growing trend is observed, it may guide the decision-making
352 (Thotakura et al. 2017). Similar considerations may be done in case of a progressive
353 reduction of ONSD when suspecting intracranial hypotension syndrome (Fichtner et al.

354 2016). In these contexts, it seems unlikely that a series of ONSD examinations to monitor
355 ICP could be performed by the same operator; in addition, the inter/intra –observer
356 variability is higher between expert and inexpert sonologists (Zeiler et al. 2013; Zeiler et
357 al. 2014). Indeed, an automated system could be extremely useful for a standardization
358 of measurements.

359 In addition, because of the variability of measurements, most ultrasonographic
360 studies use the averaging of at least two values (frequently three) to obtain the reference
361 of the ONSD, thus extending the execution time. Instead, due to the absence of
362 measurement variability when using AUTONoMA, we speculate that a single (well-
363 acquired) image could be sufficient for the ONSD measurement by using the automated
364 system, with a reduction of calculation time in comparison with the manual evaluation
365 (approximately 3-5 minutes). We specify that the automated system was able to correctly
366 segment images that presented a certain amount of variability in appearance and
367 direction of the optic nerve. In 4 images, the automated algorithm was unable to correctly
368 segment the ON and ONS due to the fact that the structures were not sufficiently
369 hypoechoic or hyperechoic (Fig. 6). Since the algorithm must make certain assumptions
370 on how specific structures are represented in the ultrasound B-mode image, if the actual
371 representation is excessively different from a typical transorbital ultrasonography image,
372 the algorithm does not properly process the image.

373 From a clinical perspective, most studies documented a 1 mm difference between
374 ONSD of healthy and pathological conditions (Lochner et al. 2017; Lochner et al. 2018b;
375 Moretti et al. 2009). Since the measured error is inferior, it is likely that the AUTONoMA
376 algorithm could distinguish between most pathological and healthy conditions.

377 An increasing number of studies examined the role of TOS for neurological
378 disorders that may affect the OND. Candelieri Merlicco et al. (Candelieri Merlicco et al.
379 2018) found that patients affected by multiple sclerosis present an atrophy of ON
380 compared to healthy subjects, and that OND values are correlated with the Kurtzke
381 Expanded Disability Status Scale (EDSS) as well as with the duration of the disease.
382 Some authors also suggested that the ultrasonographic assessment of the OND could be
383 potentially used as a biomarker for the detection of early disability in relapsing-remitting
384 multiple sclerosis (Koraysha et al. 2019). Because AUTONoMA was able to correctly
385 detect also the OND, analogue considerations of the potential role of an automated
386 measurements system can be extended for this parameter.

387 This study presents some limitations. An analysis on a larger number of images is
388 mandatory to further validate the method. To correctly process the image, AUTONoMA
389 required a substantial difference between hyperechoic and hypoechoic structures: in case
390 of insufficient quality, the automated system is not able to recognize the optic nerve.
391 Further efforts are needed to improve this algorithm in order to recognize the boundary
392 between the hypoechogenic and hyperechogenic structure of the nerve. However, it is
393 important to point out that images with a very low quality should also be excluded from a
394 manual evaluation.

395 Then, AUTONoMA was tested on images obtained from a single ultrasound
396 machine and we have no data from different ultrasound machines. Finally, a sub-analysis
397 of our data showed that AUTONoMA tended to underestimate the OND measurements
398 compared to both operators, while no conclusive information can be deducted for the
399 ONSD. We suggest that further observations are warranted in order to clarify this issue.

400 Despite these limits, our preliminary data are encouraging and can justify the use of
401 AUTONoMA as a non-invasive tool for the assessment of ONSD and OND. It is important
402 to also point out that, in the present study, the true ONSD and OND values are not known
403 and are estimated by manual measurements, which are considered as ground truth. A
404 phantom study with known OND and ONSD values would help confirm the algorithm
405 accuracy and results even further; however, this is outside the scope of the present study,
406 which aims to present a tool that can automatically measure the OND and ONSD as
407 would be done by a manual expert on a B-Mode ultrasound image.

408 In order to improve the AUTONoMA system, further investigations will be object of our
409 future studies.

410

411 **Conclusion**

412 A novel CAD system to automatically measure the OND and ONSD in ultrasound
413 images is presented. The algorithm is based on initially locating the optic bulb and optic
414 nerve centerline and then two dual snake models are implemented for the final nerve and
415 sheath segmentation. The technique was validated on a database of 71 images by
416 comparing the results with two manual operators (an expert and a non-expert operator).
417 We obtained a low mean measurement error and showed automatic results that can be
418 considered within the range of inter-operator variability. The developed system can help
419 clinicians evaluate pathologies related to the variations of the optic nerve morphology in
420 a short time and mitigate the issue of inter-operator variability. In the future, we plan on
421 testing the presented technique on a larger database to further validate the developed
422 AUTONoMA system.

423

424 **Conflict of interest:** The authors declare that there is no conflict of interest regarding the
425 publication of this article.

426

427 **Funding:** The authors received no specific funding for this work.

428 **References**

- 429 Ballantyne S., O'Neill G, Hamilton R, Hollman A. Observer variation in the sonographic
430 measurement of optic nerve sheath diameter in normal adults. *Eur J Ultrasound*
431 Elsevier, 2002;15:145–149.
- 432 Bäuerle J, Lochner P, Kaps M, Nedelmann M. Intra- and Interobserver Reliability of
433 Sonographic Assessment of the Optic Nerve Sheath Diameter in Healthy Adults. *J*
434 *Neuroimaging* 2012;22:42–45.
- 435 Bäuerle J, Niesen W-D, Egger K, Buttler K-J, Reinhard M. Enlarged Optic Nerve Sheath
436 in Aneurysmal Subarachnoid Hemorrhage despite Normal Intracranial Pressure. *J*
437 *Neuroimaging* John Wiley & Sons, Ltd (10.1111), 2016;26:194–196.
- 438 Bloria S, Bloria P, Luthra A. Is it the time to standardize the procedure of ultrasound
439 guided optic nerve sheath diameter measurement? *Saudi J Anaesth Medknow*
440 *Publications and Media Pvt. Ltd.*, 2019;13:255.
- 441 Candelieri Merlicco A, Gabaldón Torres L, Villaverde González R, Fernández Romero
442 I, Aparicio Castro E, Lastres Arias MC. Transorbital ultrasonography for measuring
443 optic nerve atrophy in multiple sclerosis. *Acta Neurol Scand* John Wiley & Sons, Ltd
444 (10.1111), 2018;138:388–393.
- 445 Caresio C, Salvi M, Molinari F, Meiburger KM, Minetto MA. Fully Automated Muscle
446 Ultrasound Analysis (MUSA): Robust and Accurate Muscle Thickness
447 Measurement. *Ultrasound Med Biol* 2017;43.
- 448 Carraro N, Servillo G, Maria Sarra V, Bignamini A, Pizzolato G, Zorzon M. Optic Nerve
449 and its Arterial-Venous Vascularization: An Ultrasonologic Study in Multiple
450 Sclerosis Patients and Healthy Controls. *J Neuroimaging* 2014;24:273–277.

451 Copetti R, Cattarossi L. Optic nerve ultrasound: artifacts and real images. *Intensive*
452 *Care Med* 2009;35:1488–1489.

453 Ertl M, Barinka F, Torka E, Altmann M, Pfister K, Helbig H, Bogdahn U, Gamulescu M,
454 Schlachetzki F. Ocular Color-Coded Sonography – A Promising Tool for
455 Neurologists and Intensive Care Physicians. *Ultraschall der Medizin - Eur J*
456 *Ultrasound* 2014;35:422–431.

457 Faita F, Gemignani V, Bianchini E, Giannarelli C, Demi M. Real-time measurement
458 system for the evaluation of the intima media thickness with a new edge detector.
459 2006 Int Conf IEEE Eng Med Biol Soc IEEE, 2006. pp. 715–718.

460 Fichtner J, Ulrich CT, Fung C, Knüppel C, Veitweber M, Jilch A, Schucht P, Ertl M,
461 Schömig B, Gralla J, Z'Graggen WJ, Bernasconi C, Mattle HP, Schlachetzki
462 F, Raabe A, Beck J. Management of spontaneous intracranial hypotension –
463 Transorbital ultrasound as discriminator. *J Neurol Neurosurg & Psychiatry* 2016;87:650 LP – 655.

465 Friedman DI, Liu GT, Digre KB. Revised diagnostic criteria for the pseudotumor cerebri
466 syndrome in adults and children. *Neurology* 2013;81:1159 LP – 1165.

467 Gerber S, Jallais M, Greer H, McCormick M, Montgomery S, Freeman B, Kane D,
468 Chittajallu D, Siekierski N, Aylward S. Automatic Estimation of the Optic Nerve
469 Sheath Diameter from Ultrasound Images BT - *Imaging for Patient-Customized*
470 *Simulations and Systems for Point-of-Care Ultrasound*. In: Cardoso MJ, Arbel T,
471 Tavares JMRS, Aylward S, Li S, Boctor E, Fichtinger G, Cleary K, Freeman B,
472 Kohli L, Shipley Kane D, Oetgen M, Pujol S, eds. Cham: Springer International
473 Publishing, 2017. pp. 113–120.

474 Goeres P, Zeiler FA, Unger B, Karakitsos D, Gillman LM. Ultrasound assessment of
475 optic nerve sheath diameter in healthy volunteers. *J Crit Care W.B. Saunders*,
476 2016;31:168–171.

477 Helmke K, Hansen HC. Fundamentals of transorbital sonographic: Evaluation of optic
478 nerve sheath expansion under intracranial hypertension. II. Patient study. *Pediatr*
479 *Radiol* 1996;26:706–710.

480 Houzé-Cerfon C-H, Bounes V, Guemon J, Le Gourrierec T, Geeraerts T. Quality And
481 Feasibility of Sonographic Measurement of the Optic Nerve Sheath Diameter to
482 Estimate the Risk of Raised Intracranial Pressure After Traumatic Brain Injury in
483 Prehospital Setting. *Prehospital Emerg Care Taylor & Francis*, 2018;1–7.

484 Koraysha NA, Kishk N, Hassan A, Samy El Gendy NM, Shehata HS, Al-Azayem SA,
485 Kamal YS. Evaluating optic nerve diameter as a possible biomarker for disability in
486 patients with multiple sclerosis. *Neuropsychiatr Dis Treat Dove Press*,
487 2019;15:2571–2578.

488 Lochner P, Behnke S, Fassbender K, Andrejewski A, Knodel S, Siniscalchi A, Jenderka
489 K-V, Lemor R. Simulation and Experimental Characterization of Lateral Imaging
490 Resolution of Ultrasound Systems and Assessment of System Suitability for
491 Acoustic Optic Nerve Sheath Diameter Measurement. *J Neuroimaging John Wiley*
492 *& Sons, Ltd* (10.1111), 2018a;0.

493 Lochner P, Cantello R, Fassbender K, Lesmeister M, Nardone R, Siniscalchi A,
494 Clemente N, Naldi A, Coppo L, Brigo F, Comi C. Longitudinal Assessment of
495 Transorbital Sonography, Visual Acuity, and Biomarkers for Inflammation and
496 Axonal Injury in Optic Neuritis. *Dis Markers Hindawi*, 2017;2017:1–5.

497 Lochner P, Coppo L, Cantello R, Nardone R, Naldi A, Leone MA, Brigo F. Intra- and
498 interobserver reliability of transorbital sonographic assessment of the optic nerve
499 sheath diameter and optic nerve diameter in healthy adults. *J Ultrasound*
500 2014;19:41–45.

501 Lochner P, Czosnyka M, Naldi A, Lyros E, Pelosi P, Mathur S, Fassbender K, Robba C.
502 Optic nerve sheath diameter: present and future perspectives for neurologists and
503 critical care physicians. *Neurol Sci Springer International Publishing*, 2019;1–11.

504 Lochner P, Falla M, Brigo F, Pohl M, Strapazzon G. Ultrasonography of the Optic Nerve
505 Sheath Diameter for Diagnosis and Monitoring of Acute Mountain Sickness: A
506 Systematic Review. *High Alt Med Biol Mary Ann Liebert, Inc., publishers*,
507 2015;16:195–203.

508 Lochner P, Fassbender K, Lesmeister M, Nardone R, Orioli A, Brigo F, Stolz E. Ocular
509 ultrasound for monitoring pseudotumor cerebri syndrome. *J Neurol* 2018b;265:356–
510 361.

511 Meiburger KM, Acharya UR, Molinari F. Automated localization and segmentation
512 techniques for B-mode ultrasound images: A review. *Comput Biol Med*
513 2018;92:210–235.

514 Molinari F, Meiburger KM, Saba L, Rajendra Acharya U, Ledda M, Nicolaides A, Suri
515 JS. Constrained snake vs. conventional snake for carotid ultrasound automated
516 IMT measurements on multi-center data sets. *Ultrasonics Elsevier B.V.*,
517 2012a;52:949–961.

518 Molinari F, Meiburger KM, Saba L, Zeng G, Acharya UR, Ledda M, Nicolaides A, Suri
519 JS. Fully automated dual-snake formulation for carotid intima-media thickness

520 measurement: A new approach. *J Ultrasound Med* 2012b;31.

521 Moretti R, Pizzi B, Cassini F, Vivaldi N. Reliability of Optic Nerve Ultrasound for the
522 Evaluation of Patients with Spontaneous Intracranial Hemorrhage. *Neurocrit Care*
523 2009;11:406–410.

524 Naldi A, Provero P, Vercelli A, Bergui M, Mazzeo AT, Cantello R, Tondo G, Lochner P.
525 Optic nerve sheath diameter asymmetry in healthy subjects and patients with
526 intracranial hypertension. *Neurol Sci Springer International Publishing*, 2019;1–5.

527 Ossoinig K. Standardized echography: basic principles, clinical applications, and
528 results. *journals.lww.com* 1979;127–210.

529 Robba C, Bacigaluppi S, Cardim D, Donnelly J, Bertuccio A, Czosnyka M. Non-invasive
530 assessment of intracranial pressure. *Acta Neurol Scand Wiley/Blackwell* (10.1111),
531 2015;134:4–21.

532 Robba C, Santori G, Czosnyka M, Corradi F, Bragazzi N, Padayachy L, Taccone FS,
533 Citerio G. Optic nerve sheath diameter measured sonographically as non-invasive
534 estimator of intracranial pressure: a systematic review and meta-analysis. *Intensive*
535 *Care Med Springer Berlin Heidelberg*, 2018;44:1284–1294.

536 Saba L, Molinari F, Meiburger K, Piga M, Zeng G, Acharya UR, Nicolaidis A, S Suri J.
537 What is the correct distance measurement metric when measuring carotid
538 ultrasound intimamedia thickness automatically? *Int Angiol* 2012;31:483–489.

539 Schroeder W. A-scan-ultrasonography in unilateral optic nerve lesions. *Klin Monbl*
540 *Augenheilkd* 1976;169:30–8.

541 Soliman I, Johnson GGRJ, Gillman LM, Zeiler FA, Faqih F, Aletreby WT, Balhamar A,
542 Mahmood NN, Ahmad Mumtaz S, Alharthy A, Lazaridis C, Karakitsos D. *New Optic*

543 Nerve Sonography Quality Criteria in the Diagnostic Evaluation of Traumatic Brain
544 Injury. Crit Care Res Pract Hindawi, 2018;2018:1–7.

545 Soroushmehr R, Rajajee K, Williamson C, Gryak J, Najarian K, Ward K, Tiba MH.
546 Automated Optic Nerve Sheath Diameter Measurement Using Super-pixel
547 Analysis. 41st Annu Int Conf IEEE Eng Med Biol Soc Institute of Electrical and
548 Electronics Engineers (IEEE), 2019. pp. 2793–2796.

549 Thotakura AK, Marabathina NR, Danaboyina AR, Mareddy RR. Role of serial ultrasonic
550 optic nerve sheath diameter monitoring in head injury. Neurochirurgie Elsevier
551 Masson, 2017;63:444–448.

552 Wang L, Feng L, Yao Y, Deng F, Wang Y, Feng J, Xing Y. Ultrasonographic Evaluation
553 of Optic Nerve Sheath Diameter among Healthy Chinese Adults. Ultrasound Med
554 Biol 2016;42:683–688.

555 Zeiler FA, Unger B, Kramer AH, Kirkpatrick AW, Gillman LM. A Unique Model for
556 Ultrasound Assessment of Optic Nerve Sheath Diameter. Can J Neurol Sci / J Can
557 des Sci Neurol Cambridge University Press, 2013;40:225–229.

558 Zeiler FA, Unger B, Zhu Q, Xiao J, Kirkpatrick AW, Kramer AH, Aleassa EM, Gillman
559 LM. A Unique Model for ONSD Part II: Inter/Intra-operator Variability. Can J Neurol
560 Sci / J Can des Sci Neurol Cambridge University Press, 2014;41:430–435.

561

562

563 **Figures Captions List**

564

565 Figure 1. Example of manual optic nerve diameter (OND) and optic nerve sheath diameter
566 (ONSD) calculation.

567

568 Figure 2. Overview of steps for the developed AUTONoMA system. The image is first
569 acquired, and then automatically cropped. Stage I consists in the automatic recognition
570 of the bulb and optic nerve centerline. Stage II then consists in the initialization of the two
571 dual snake models by a rough segmentation of the optic nerve and optic nerve sheath.
572 The dual snake models then evolve in time until they reach the borders of the actual optic
573 nerve and optic nerve sheath. Then the final value of the OND and ONSD is automatically
574 measured from the final dual snake boundaries.

575

576 Figure 3. Overview of the AUTONoMA Stage I architecture. A) Original image. B)
577 Automatically cropped image. C) First order gaussian derivative of (b), showing the
578 initialization of the search region for the bulb profile tracing. D) AUTONoMA bulb
579 ($AUTONoMA_{bulb}$)profile segmentation results. E) AUTONoMA optic nerve centerline
580 ($AUTONoMA_{onc}$) tracing.

581

582 Figure 4. Overview of the AUTONoMA Stage II architecture. A) Optic nerve (ON) dual
583 snake initialization. B) Optic nerve sheath (ONS) dual snake initializaion. C) Final ON dual
584 snake segmentations ($AUTONoMA_{ONL}$ and $AUTONoMA_{ONR}$). D) Final ONS dual snake

585 segmentations (AUTONoMA_{ONSL} and AUTONoMA_{ONSR}). E) Final calculation of the
586 automatic OND and ONSD measurements (AUTONoMA_{OND} and AUTONoMA_{ONSD}).

587

588 Figure 5. Segmentation and OND and ONSD measurement results of the developed
589 AUTONoMA system.

590

591 Figure 6. Example error cases for the AUTONoMA system. A) Example where there is
592 not a sufficient intensity difference between the hypoechogenicity of the optic nerve and
593 the surrounding arachnoid space. B) Example of an image where the surrounding
594 arachnoid space is excessively hyperechoic.

595

596 Figure 7a. Bland-Altman analysis comparing the optic nerve diameter (OND – first row)
597 and the optic nerve sheath diameter (ONSD – second row) with Operator 1 (first column)
598 and Operator 2 (second column). Continuous line depicts the mean of differences;
599 dashed lines denote limits of agreement.

600

601 Figure 7b. Bland-Altman inter-operator analysis comparing the optic nerve diameter
602 (OND – first column) and the optic nerve sheath diameter (ONSD – second column) with
603 Operator 1 vs Operator 2. Continuous line depicts the mean of differences; dashed lines
604 denote limits of agreement.

605

606

607

608 **Tables**

609

610 Table 1

611 Parameter values for Dual Snake models. ONDS: Optic Nerve Dual Snake; ONSDS:
612 Optic Nerve Sheath Dual Snake

Dual Snake model	α (internal energy)	β (external energy)	γ (mutual energy)
ONDS	0.7	$\frac{0.3 \cdot CF_{base}}{CF}$	$\frac{9 \cdot CF_{base}}{CF}$
ONSDS	0.3	$\frac{0.1 \cdot CF_{base}}{CF}$	$\frac{6 \cdot CF_{base}}{CF}$

613

614 Table 2

615 Performance evaluation results of the AUTONoMA system compared to manual
616 measurements for the calculation of the optic nerve diameter (OND).

OND	AUTONoMA	Operator 1	Operator 2
Mean value [mm]	3.1 ± 0.3	3.1 ± 0.4	3.0 ± 0.4
Mean error [mm]		0.06 ± 0.35	-0.05 ± 0.38
Mean absolute error [mm]		0.28 ± 0.22	0.30 ± 0.24
Mean squared error [mm ²]		0.12 ± 0.17	0.15 ± 0.25
FoM		98.2%	98.3%

617

618 Table 3

619 Performance evaluation results of the AUTONoMA system compared to manual
620 measurements for the calculation of the optic nerve sheath diameter (ONSD).

ONSD	AUTONoMA	Operator 1	Operator 2
Mean value [mm]	6.2 ± 0.6	6.2 ± 0.6	5.8 ± 0.6*
Mean error [mm]		0.06 ± 0.52	-0.37 ± 0.55
Mean absolute error [mm]		0.41 ± 0.32	0.49 ± 0.45
Mean squared error [mm ²]		0.27 ± 0.37	0.44 ± 0.66
FoM		99.0%	93.5%

*statistically significant difference (p<0.05) between Operator and AUTONoMA using the Wilcoxon signed rank test

621

622

623

624 Table 4

625 Correlation performance results between the AUTONoMA system and the manual
626 operators (Op1 and Op2) and the inter-operator variability performance analysis.

Analysis	Measure	Correlation coefficient	Confidence interval		p-value
			Lower limit	Upper limit	
AUTONoM	OND	0.47	0.27	0.63	3.590 · 10 ⁻⁵
A vs Op1	ONSD	0.64	0.48	0.76	1.541 · 10 ⁻⁹
AUTONoM	OND	0.35	0.12	0.54	0.0031
A vs Op2	ONSD	0.61	0.44	0.74	1.375 · 10 ⁻⁸
Op1 vs	OND	0.69	0.55	0.80	2.543 · 10 ⁻¹¹
Op2	ONSD	0.65	0.49	0.77	7.222 · 10 ⁻¹⁰

627