

Novel Load Systems for In Vitro Testing of Biomaterials and Medical Devices

Original

Novel Load Systems for In Vitro Testing of Biomaterials and Medical Devices / Bignardi, Cristina; Terzini, Mara. - In: MATERIALS. - ISSN 1996-1944. - 16:2(2023), p. 465. [10.3390/ma16020465]

Availability:

This version is available at: 11583/2975658 since: 2023-02-06T09:42:13Z

Publisher:

MDPI

Published

DOI:10.3390/ma16020465

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

Editorial

Novel Load Systems for *In Vitro* Testing of Biomaterials and Medical Devices

Cristina Bignardi ^{1,2,*}  and Mara Terzini ^{1,2} ¹ Department of Mechanical and Aerospace Engineering, Politecnico di Torino, 10129 Turin, Italy² Polito^{BIO}Med Lab, Politecnico di Torino, 10129 Turin, Italy

* Correspondence: cristina.bignardi@polito.it

In the mechanical characterization of materials or devices, the real load conditions to which they will be subjected in their operational environment must often be simulated by starting from the availability of universal testing machines. Although sophisticated, these machines only apply uniaxial loads or, at most, simultaneously apply loads along few degrees of freedom.

With the aim of investigating the structural behavior of materials and devices, usually researchers check the availability of adequate existing testing machines before inventing new loading systems and verify the existence of indications in the International Standards or in the literature. In the field of biomedical engineering complexity grows due to the peculiar loads (i.e., the physiological ones) and environments to which biomaterials and medical devices are subjected. They, indeed, will be coupled to the receiving organs or tissues once implanted.

This Special Issue was born from the desire to give space to the imagination of the researchers involved in the conception and design of experimental benches able to simulate the real load conditions under which biomaterials and medical devices will operate. In the Special Issue, seven papers were published which explore different contexts in which the design of novel loading systems constituted the originality of the research. They range from the characterization of the mechanical behavior of biomaterials [1,2] and of biological tissues [3], to the verification of the structural behavior of more or less complex medical devices [4,5], from the design of tests for verifying the stability of articular prosthetic components once coupled to the bone [6], to the simulation of the erosive effect of surgical meshes on the biological soft tissues with which they interface [7].

The aim of the investigations [1,2], in the ambit of implant fixed dental prostheses, is to determine the influence of luting agent, abutment height and taper angle on the retrievability of abutment-coping cementations in two different layouts: with one implant and with two implants connected by a bridge. The number of impulses and the impulsive force delivered during each test were recorded and used as retrievability indexes.

Only recently, ovary responsiveness to mechanical signals was exploited for reproductive purposes. Poor characterization of ovarian cortex biomechanics and of the mechanical challenge hampers reproducible and effective treatments, and prevention of tissue damages. In [3], the biomechanical response of ovarian cortical tissue from abattoir bovines is characterized for the first time. Ovarian cortical tissue fragments were subjected to uniaxial dynamic testing at frequencies up to 30 Hz, and at increasing average stresses. In [4], both mechanical and morphological properties of an innovative bandage system, useful to position and fix patients' torsos or extremities, are reported. The creation of a new complex medical device, such as a prosthesis for amputees, requires numerous design phases accompanied by fundamental structural tests. In [5], all the design phases of a new powered ankle-foot prosthesis are reported. Sufficient primary fixation stability is the basis for the osseointegration of cementless acetabular cups. Several test methods have been established for determining the tilting moment of acetabular press-fit cups, which is a measure for their



Citation: Bignardi, C.; Terzini, M. Novel Load Systems for *In Vitro* Testing of Biomaterials and Medical Devices. *Materials* **2023**, *16*, 465. <https://doi.org/10.3390/ma16020465>

Received: 20 December 2022

Accepted: 30 December 2022

Published: 4 January 2023



Copyright: © 2023 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

primary fixation stability. The central aim of the experimental study explained in [6] is to show the differences between the commonly used lever-out test method and the edge-load test method in which the cup insert is axially loaded during the tilting process with respect to the parameters, tilting moment, and interface stiffness. At last, the research shown in [7] deals with the phenomenon of meshes erosion whereby soft tissue becomes damaged as a result of contact with implants made from surgical mesh, a fabric-like material consisting of fibers of polypropylene or other polymers. The paper describes the design and construction of a testing machine to generate mesh erosion *in vitro*.

The Editors give special thanks to the authors and the editorial team of *Materials* for the collaborative and peer-review process. We hope that the scientific community of *Materials* will enjoy reading this Special Issue and find new ideas for future and present research works.

Author Contributions: Conceptualization, C.B. and M.T.; writing—original draft preparation, C.B. and M.T.; writing—review and editing, C.B. and M.T.; supervision, C.B. and M.T. All authors have read and agreed to the published version of the manuscript.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Lugas, A.T.; Terzini, M.; Zanetti, E.M.; Schierano, G.; Manzella, C.; Baldi, D.; Bignardi, C.; Audenino, A.L. In Vitro Simulation of Dental Implant Bridges Removal: Influence of Luting Agent and Abutments Geometry on Retrieval. *Materials* **2020**, *13*, 2797. [[CrossRef](#)] [[PubMed](#)]
2. Lugas, A.T.; Terzini, M.; Zanetti, E.M.; Schierano, G.; Manzella, C.; Baldi, D.; Bignardi, C.; Audenino, A.L. In Vitro Impact Testing to Simulate Implant-Supported Prosthesis Retrieval in Clinical Practice: Influence of Cement and Abutment Geometry. *Materials* **2020**, *13*, 1749. [[CrossRef](#)] [[PubMed](#)]
3. Pascoletti, G.; Di Nardo, M.; Fragomeni, G.; Barbato, V.; Capriglione, T.; Gualtieri, R.; Talevi, R.; Catapano, G.; Zanetti, E.M. Dynamic Characterization of the Biomechanical Behaviour of Bovine Ovarian Cortical Tissue and Its Short-Term Effect on Ovarian Tissue and Follicles. *Materials* **2020**, *13*, 3759. [[CrossRef](#)] [[PubMed](#)]
4. Alleva, S.; Antonelli, M.G.; Beomonte, Z.P.; Durante, F. Biomechanical Design and Prototyping of a Powered Ankle-Foot Prosthesis. *Materials* **2020**, *13*, 5806. [[CrossRef](#)] [[PubMed](#)]
5. Putzer, D.; Dammerer, D.; Coraça-Huber, D.; Pallua, J.; Schmölz, W.; Nogler, M. Mechanical and Morphological Assessment of an Innovative Textile for Patient Positioning Applications: Comparison to Two Standard Bandage Systems. *Materials* **2021**, *14*, 1508. [[CrossRef](#)] [[PubMed](#)]
6. Schulze, C.; Vogel, D.; Mallow, S.; Bader, R. Comparison of Test Setups for the Experimental Evaluation of the Primary Fixation Stability of Acetabular Cups. *Materials* **2020**, *13*, 3982. [[CrossRef](#)] [[PubMed](#)]
7. Schmidt, A.; O'Brien, G.; Taylor, D. The Development of Equipment to Measure Mesh Erosion of Soft Tissue. *Materials* **2021**, *14*, 941. [[CrossRef](#)] [[PubMed](#)]

Disclaimer/Publisher's Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.