

Estimation of muscle fiber conduction velocity in pinnate muscles: a simulation study

*Original*

Estimation of muscle fiber conduction velocity in pinnate muscles: a simulation study / Mesin, Luca; Farina, D.. - (2006). (XVI Congresso della International Society of Electrophysiology & Kinesiology (ISEK) Torino 29 giugno – 1° luglio 2006.).

*Availability:*

This version is available at: 11583/1919684 since:

*Publisher:*

*Published*

DOI:

*Terms of use:*

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

(Article begins on next page)

## T01: EMG modeling

### ESTIMATION OF MUSCLE FIBER CONDUCTION VELOCITY IN PINNATE MUSCLES: A SIMULATION STUDY


Mesin L.<sup>(1)</sup>, Farina D.<sup>(2)</sup>

<sup>(1)</sup> LISiN, Dept. of Electronics, Politecnico di Torino, Torino, Italy. <sup>(2)</sup> Center for Sensory-Motor Interaction (SMI), Department of Health Science and Technology, Aalborg University, Aalborg, Denmark

**AIMS:** The aim of this simulation study was to assess the feasibility of estimating muscle fibre conduction velocity (CV) from surface electromyographic (EMG) signals in muscles with one and two pinnation angles.

**METHODS:** The volume conductor consisted of a layered medium simulating anisotropic muscle tissue and isotropic homogeneous subcutaneous tissue. The muscle tissue was homogeneous in case of one pinnation angle and inhomogeneous for bipinnate muscles (two fibre directions). Interference EMG signals were obtained by simulating recruitment thresholds and discharge patterns of a set of 100 or 200 motor units for the pinnate and bipinnate muscle, respectively. CV was estimated from two surface EMG channels in case of monopolar, single differential, double differential, and Laplacian recordings. The simulated CV value was in all cases 4 m/s.

**RESULTS:** Estimates of CV from the pinnate and bipinnate muscle are shown in the Figure. The bias in CV was on average larger than 0.8 m/s in all cases.



**CONCLUSIONS:** The main factors affecting CV estimates were the end-plate and end-of-fibre components due to the scatter of the projection of end-plate and tendon locations along the fibre direction, as a consequence of pinnation. The same problem

arises for muscles with the line of innervation zone locations not perpendicular to fibre direction. These results provide an interpretation of the positive bias in CV estimates from surface EMG signals, usually observed in experimental conditions.