

Prediction of cyclic fatigue life of NiTi rotary files by Virtual Modeling and Finite Elements Analysis

*Original*

Prediction of cyclic fatigue life of NiTi rotary files by Virtual Modeling and Finite Elements Analysis / Scattina, Alessandro; Alovise, Mario; Paolino, Davide Salvatore; Pasqualini, Damiano; Scotti, Nicola; Chiandussi, Giorgio; Berutti, Elio. - In: JOURNAL OF ENDODONTICS. - ISSN 0099-2399. - 41:11(2015), pp. 1867-1870. [10.1016/j.joen.2015.07.010]

*Availability:*

This version is available at: 11583/2620751 since: 2015-10-28T07:35:08Z

*Publisher:*

Elsevier Inc.

*Published*

DOI:10.1016/j.joen.2015.07.010

*Terms of use:*

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

(Article begins on next page)

**Prediction of cyclic fatigue life of NiTi rotary files by Virtual Modeling and Finite Elements Analysis**

A. Scattina<sup>2</sup>, M. Alovise<sup>1</sup>, D. Paolino<sup>2</sup>, D. Pasqualini<sup>1</sup>, N. Scotti<sup>1</sup>, G. Chiandussi<sup>2</sup>, E. Berutti<sup>1</sup>

1 University of Turin, Department of Surgical Sciences, Dental School, Endodontics, Turin, Italy

2 Department of Mechanical and Aerospace Engineering, Politecnico di Torino, Turin, Italy

## **Abstract**

**Introduction:** Finite element method (FEM) has been proposed as a method to analyze stress distribution in nickel-titanium (NiTi) rotary instruments, but has not been assessed as a method of predicting the number of cycles to failure (NCF). The objective of this study was to predict NCF and failure location of NiTi rotary instruments by FEM virtual simulation of an experimental non static fatigue test. **Material and methods:** ProTaper Next (PTN) X1, X2, X3 files (N = 20 each) were tested to failure using a customized fatigue testing device. The device and file geometries were replicated with computer aided design (CAD) software. CAD geometries (geometrical model) were imported and discretized (numerical model). The typical material model of an M-Wire alloy was applied. The numerical model of the device and file geometries were exported for finite element analysis (FEA). Multiaxial random fatigue methodology was used to analyze stress history and predict instrument life. Experimental data from PTN X2 and X3 were utilized for virtual model tuning through a reverse engineering approach to optimize material mechanical properties. Tuned material parameters were used to predict the average NCF and failure locations of PTN X1 by FEA. T-tests were used to compare FEA and experimental findings ( $P < .05$ ). **Results:** Experimental NCF and failure locations did not differ from those predicted with FEA ( $P = .098$ ). **Conclusion:** File NCF and failure location may be predicted by FEA. Virtual design, testing and analysis of file geometries could save considerable time and resources during instrument development.

## **Key words**

C-S criterion; finite element analysis; multiaxial random fatigue; NiTi rotary instrument; virtual modeling

## **Introduction**

Since their introduction, nickel-titanium (NiTi) rotary instruments have considerably simplified endodontic practice by improving speed, accuracy and safety during root canal shaping (1, 2). Continuous enhancements in the design and manufacture of NiTi rotary instruments have significantly reduced, but not completely eliminated, the occurrence of failure during canal shaping (3, 4). Although rare, breakage of NiTi rotary instruments may lead to incomplete chemomechanical treatment of the root canal system and hinder healing, particularly in the presence of pre-existing apical radiolucency (5, 6). Furthermore excessive removal of the tooth structure can occur during fragment elimination procedures (5, 6). While many variables can contribute to instrument failure, the primary causes are cyclic bending fatigue and torsional overload. Torsional overload is related to the risk of taper lock and its incidence might be reduced by performing glide path and preliminary enlargement (7–9). However, NiTi instrument failures are mainly due to cyclic bending fatigue, which occurs when a NiTi instrument rotates in a curved root canal (10–14). During rotation, the instrument material is alternatively subjected to compressive and tensile stresses (11). Such stresses may initiate crack formation and propagation, leading to eventual failure (4, 15). Instruments that fail due to bending fatigue usually exhibit no specific macroscopic patterns and failure may occur without any visible warning (1, 5, 16, 17). Canal shape, instrument geometry, rotational speed, torque, instrument surface treatments and the chemical composition of NiTi alloys are the main factors affecting the number of cycles to failure (NCF) of NiTi rotary instruments (2, 4, 18-21). Standardized experimental conditions are not possible for extracted teeth (22), but several self-designed devices and methods have been used to assess NCF of NiTi rotary instruments in vitro (22). However, there is no international standard for testing the cyclic fatigue behavior of endodontic rotary instruments in vitro (22). A three-dimensional computerized approach based on the finite element method (FEM) (23–27) has been recently proposed to analyze stress distribution

in bending fatigue. This approach takes cyclic loading conditions into consideration (25, 27), but makes no attempt to predict NCF.

The objective of this study was to assess the possibility of predicting the NCF and failure location of NiTi rotary instruments by the virtual simulation of an experimental fatigue test (non-static loading condition) by the FEM.

## **Materials and methods**

ProTaper Next (PTN) X1, X2 and X3 files (N = 20 each), 25 mm in length, were tested to failure using a tempered steel fatigue testing device developed by Dentsply Maillefer Laboratories (Baillagues, Switzerland), which produced a reproducible simulation of rotary instrumentation within a curved canal. Testing configurations adopted for every file type are reported in Figure 1 and Table 1. Testing configurations for file X1 and files X2 and X3 differ with respect to canal depth and the position of the curvature center of the component simulating the inner canal wall. The testing device is covered with a glass or steel element in order to prevent the file leaving the canal during rotation. Each file was inserted leaving the non-cutting tip outside the canal and rotation initiated with an electric motor (X-Smart Dentsply Maillefer) at the suggested settings (300 rpm, 5 Ncm). Time to failure was recorded and converted to NCF. Experimental tests were carried out by Dentsply Maillefer Laboratories: tested files were photographed with a digital camera after failure (Canon EOS 350D, 8 Mpx resolution, ISO 100, f 18, 1/60 s) and failure locations were measured in millimeters from the tip of the instrument (image analysis, caliper).

The customized fatigue testing device and file geometries were computer-replicated (*videoclip virtual device.mov*) with Computer Aided Design (CAD) software (SolidWorks 2013, Dassault Systems, Waltham, Massachusetts, USA). CAD geometries (geometrical model) were imported in a pre-post processor (Altair Hypermesh 12.0, Troy, MI, USA) and discretized (numerical model). The

main body of each file was modeled with 18,000 brick elements, with an average size of 0.1 mm (Fig. 2). The tip and neck of the file were modeled with tetrahedral elements with the same average dimensions. The typical material model of an M-Wire alloy (NiTi alloy), taken from the literature (28), was applied. The testing device was modeled with 2000 shell element and was considered a rigid material. Contact between the file and the testing device was introduced with a penalty contact formulation implemented in LS-Dyna (LS-Dyna R7.1, Livermore, USA). The penalty method introduces a force at the contact detection points that has penetrated across the target surface with the express purpose of eliminating the penetration (29). The numerical model of the customized fatigue testing device and the file geometries were exported into the finite element (FE) solver (LS-Dyna R7.1, Livermore, USA) for finite element analyses (FEA). The FEA aimed to mimic the experimental test carried out with the Maillefer customized fatigue testing device. The file, in rest configuration (straight file), was inserted into the testing device and then bent by the simulated canal walls (rigid shell elements). It was finally put into rotation at 300 rpm. The stress state in the brick elements of each file model was recorded every 0.2 s for 2 s (*videoclip: X1\_stress.mov, X2\_stress.mov, X3\_stress.mov*).

Traditional life prediction methods developed for uniaxial variable amplitude loading could not be used in this study due to the complexity of the stress-strain state and the loading history over time (multiaxial random loading). Element stress history must be analyzed by applying an appropriate multiaxial random fatigue (MRF) criterion. Few methodologies concerning MRF have been proposed and research on the topic is still in progress. A recent MRF methodology, proposed by Carpinteri et al. (30) and named the C-S criterion, was adopted in this study in order to analyze the stress history of each file element. The C-S criterion was implemented in Matlab 2014 (The MathWorks Inc, Natick, MA) with the assumption that the stress history recorded in the first 2 s was identically repeated until failure. The C-S criterion was used to calculate the time to failure for each brick element and, as a consequence, to predict the file NCF (the minimum NCF in the file).

The mechanical properties of a material are not provided and the influence of manufacturing surface defects cannot be evaluated a priori. As a consequence, a reverse engineering approach for model tuning was adopted to obtain the required information. Several material properties were taken into account through the Matlab model and were tuned to match the experimental NCF of PTN X2 and X3 files. Material parameters were altered through an optimization algorithm implemented in Matlab 2014 in order to reduce the difference between the file NCF obtained with FEM and the experimental NCF. The tuned material parameters were then used to predict the average NCF of PTN X1 file with FEA. Failure locations for each file type were recorded and compared with the failure locations of the files photographed after failure in the Maillefer customized fatigue testing device. T-tests were performed to compare FEA and experimental results ( $P = .05$ ).

## **Results**

The main results of the analyses are presented in Table 2.

NCF of PTN X2 and X3 files was used for the tuning of material parameters involved in the C-S criterion. Their P-values for the NCF were subsequently extremely large ( $P = .896$  and  $.787$ , respectively). There was also no significance in the NCF of the PTN X1 file ( $P = .098$ ). Therefore, no significant statistical difference was found between the experimental NCF and the NCF predicted by FEA.

In addition, no significant statistical difference was found between the experimental failure locations and the failure locations predicted by FEA ( $P = .148$  for PTN X1 file,  $P = .995$  for PTN X2 file and  $P = .481$  for PTN X3 file).

Possible failure locations predicted with FEA, together with the experimental failure locations obtained from the photographed files, are shown in Figure 3. For every file type, vertical lines

indicate 95% confidence intervals of the experimental failure locations, while the highlighted point denotes the location of the element with the smallest NCF where the predicted failure occurred.

## **Discussion**

Multiaxial loading usually acts on NiTi rotary instruments during shaping (31, 32) as a result of the simultaneous presence of bending and torsion. Given a file type (geometry and material), the torsional overload largely depends on the pre-existing root canal size and the apical force exerted during instrumentation, while the cyclic bending fatigue depends on un-modifiable anatomic factors such as the root canal radius and degree of curvature (22). File rotation around a curved axis (the axis of the file inserted into a curved canal root) leads to a continuously varying complex stress state including normal and shear components (15) and therefore to a fatigue phenomenon, which accounts for up to 90% of endodontic instrument mechanical failures (2, 33).

In the present study, FEA analysis was introduced as a virtual testing approach to predict mechanical behavior of the NiTi file. FEA is a numerical method that calculates the stress field in any file geometry, according to the file material properties and the boundary loading conditions (27). Virtual testing results obtained by simulating the rotation of ProTaper Next X1, X2 and X3 instruments were compared to the experimental results obtained with a customized fatigue testing device. A reverse engineering approach was utilized to overcome the lack of information regarding the material behavior and the local stress concentration effects caused by manufacturing surface defects (5, 34, 35).

The instrument stress state obtained by FEA was elaborated according to the C-S criterion proposed by Carpinteri et al. (30) in order to calculate the NCF and failure location due to cyclic bending fatigue. The C-S criterion is applicable when stresses are below the yield strength of the material (i.e. no plasticity phenomena occur during load application). The applicability of the C-S- criterion

is not influenced by file rotation speed or, more generally, by the applied load frequency. The Authors verified the applicability of the C-S criterion: for the analysed cases, the stress state was far below the yield strength of the material and therefore no plasticity occurred during load application.

Unknown material properties (e.g. the normal stress fatigue limit and the shear stress fatigue limit) were required for the application of the C-S criterion and a reverse engineering approach was carried out in order to estimate them. Material properties were modified and optimized to reduce the difference between the file NCF obtained with FEA and the NCF obtained with the fatigue testing device for ProTaper Next X2 and X3. The tuned parameters are specific for the material used to manufacture the analyzed files. The same parameters can be utilized for any FEA in which the file material is kept the same. If the file material changes, the material properties change and must be found through a further tuning process. The optimized material properties and the fatigue prediction model based on the C-S criterion were used to predict the NCF of the ProTaper Next X1.

The results presented here demonstrate that virtual simulation through FEA can be usefully adopted to predict the file NCF and the failure location. As a consequence, given the material properties obtained through the reverse engineering approach, FEA can be effectively used to virtually design new file geometries. Therefore, the FEA approach would potentially overcome the usual expensive trial and error approach based on repetitive prototype manufacturing and testing. This study demonstrates the feasibility of numerically analysing instrument geometries to predict the file NCF and location of failure. The evaluation of material mechanical properties through a reverse engineering approach based on a multiaxial random fatigue criterion allows the virtual verification of new file geometries toward cyclic bending fatigue failure. The virtual analysis of file behavior may lead to a considerable saving of time and resources during instrument development.

The evaluation of the material properties through a reverse engineering approach could be avoided if the material properties and the manufacturing defects are properly characterized and taken into account.

## References

1. Al-Sudani D, Grande NM, Plotino G, et al. Cyclic fatigue of nickel-titanium rotary instruments in a double (S-shaped) simulated curvature. *J Endod* 2012;38:987–9.
2. Bhagabati N, Yadav S, Talwar S. An in vitro cyclic fatigue analysis of different endodontic nickel-titanium rotary instruments. *J Endod* 2012;38:515–8.
3. Li UM, Lee BS, Shih CT, Lan WH, Lin CP. Cyclic fatigue of endodontic nickel titanium rotary instruments: static and dynamic tests. *J Endod* 2002;28:448–51.
4. Lopes HP, Gambarra-Soares T, Elias CN, et al. Comparison of the mechanical properties of rotary instruments made of conventional nickel-titanium wire, M-wire, or nickel-titanium alloy in R-phase. *J Endod* 2013;39:516–20.
5. Anderson ME, Price JW, Parashos P. Fracture resistance of electropolished rotary nickel-titanium endodontic instruments. *J Endod* 2007;33:1212–6.
6. Spili P, Parashos P, Messer HH. The impact of instrument fracture on outcome of endodontic treatment. *J Endod* 2005;31:845–50.
7. Berutti E, Negro AR, Lendini M, Pasqualini D. Influence of manual preflaring and torque on the failure rate of ProTaper rotary instruments. *J Endod* 2004;30:228–30.
8. Berutti E, Cantatore G, Castellucci A, et al. Use of nickel-titanium rotary PathFile to create the glide path: comparison with manual preflaring in simulated root canals. *J Endod* 2009;35:408–12.
9. Patiño PV, Biedma BM, Liébana CR, Cantatore G, Bahillo JG. The influence of a manual glide path on the separation rate of NiTi rotary instruments. *J Endod* 2005;31:114–6.
10. Sattapan B, Nervo GJ, Palamara JE, Messer HH. Defects in rotary nickel-titanium files after clinical use. *J Endod* 2000;26:161–5.
11. Lopes HP, Ferreira AA, Elias CN, Moreira EJ, de Oliveira JC, Siqueira JF Jr. Influence of

rotational speed on the cyclic fatigue of rotary nickel-titanium endodontic instruments. *J Endod* 2009;35:1013–6.

12. Parashos P, Gordon I, Messer HH. Factors influencing defects of rotary nickel-titanium endodontic instruments after clinical use. *J Endod* 2004;30:722–5.

13. Cheung GS, Peng B, Bian Z, Shen Y, Darvell BW. Defects in ProTaper S1 instruments after clinical use: fractographic examination. *Int Endod J* 2005;38:802–9.

14. Peng B, Shen Y, Cheung GS, Xia TJ. Defects in ProTaper S1 instruments after clinical use: longitudinal examination. *Int Endod J* 2005;38:550–7.

15. Bouska J, Justman B, Williamson A, DeLong C, Qian F. Resistance to cyclic fatigue failure of a new endodontic rotary file. *J Endod* 2012;38:667–9.

16. Yum J, Cheung GS, Park JK, Hur B, Kim HC. Torsional strength and toughness of nickel-titanium rotary files. *J Endod* 2011;37:382–6.

17. Pirani C, Cirulli PP, Chersoni S, Micele L, Ruggeri O, Prati C. Cyclic fatigue testing and metallographic analysis of nickel-titanium rotary instruments. *J Endod* 2011;37:1013–6.

18. Elnaghy AM, Elsaka SE. Assessment of the Mechanical Properties of ProTaper Next Nickel-Titanium Rotary Files. *J Endod* 2014;40:1830–4.

19. Shen Y, Qian W, Abtin H, Gao Y, Haapasalo M. Fatigue testing of controlled memory wire nickel-titanium rotary instruments. *J Endod* 2011;37:997–1001.

20. Gao Y, Gutmann JL, Wilkinson K, Maxwell R, Ammon D. Evaluation of the impact of raw materials on the fatigue and mechanical properties of ProFile Vortex rotary instruments. *J Endod* 2012;38:398–401.

21. Johnson E, Lloyd A, Kuttler S, Namerow K. Comparison between a novel nickel-titanium alloy and 508 nitinol on the cyclic fatigue life of ProFile 25/.04 rotary instruments. *J Endod* 2008;34:1406–9.

22. Plotino G, Grande NM, Cordaro M, Testarelli L, Gambarini G. A review of cyclic fatigue testing of nickel-titanium rotary instruments. *J Endod* 2009;35:1469–76.
23. Berutti E, Chiandussi G, Gaviglio I, Ibba A. Comparative analysis of torsional and bending stresses in two mathematical models of nickel-titanium rotary instruments: ProTaper versus ProFile. *J Endod* 2003;29:15–9.
24. Xu X, Eng M, Zheng Y, Eng D. Comparative study of torsional and bending properties for six models of nickel–titanium root canal instruments with different cross-sections. *J Endod* 2006;32:372–5.
25. Kim HC, Kim HJ, Lee CJ, Kim BM, Park JK, Versluis A. Mechanical response of nickel-titanium instruments with different cross-sectional designs during shaping of simulated curved canals. *Int Endod J* 2009;42:593–602.
26. Montalvão D, Shengwen Q, Freitas M. A study on the influence of Ni-Ti M-Wire in the flexural fatigue life of endodontic rotary files by using Finite Element Analysis. *Mater Sci Eng C Mater Biol Appl* 2014;40:172–9.
27. Lee MH, Versluis A, Kim BM, Lee CJ, Hur B, Kim HC. Correlation between experimental cyclic fatigue resistance and numerical stress analysis for nickel-titanium rotary files. *J Endod* 2011;37:1152–7.
28. Auricchio F, Taylor RL, Lubliner J. Shape-memory alloys: macromodelling and numerical simulations of the superelastic behavior. *Comput. Methods Appl. Mech. Engng.* 1997;146:281–312.
29. LS-DYNA®. Theory Manual. Livermore Software Technology Corporation (LSTC). Livermore, California.
30. Carpinteri A, Spagnoli A, Vantadori S, Bagni C. Structural integrity assessment of metallic components under multiaxial fatigue: the C–S criterion and its evolution. *Fatigue Fract Engng Mater Struct* 2013;36:870–83.

31. Bahia MG, Melo MC, Buono VT. Influence of cyclic torsional loading on the fatigue resistance of K3 instruments. *Int Endod J* 2008;41:883–91.
32. Bahia MG, Martins RC, Gonzalez BM, Buono VT. Physical and mechanical characterization and the influence of cyclic loading on the behaviour of nickel-titanium wires employed in the manufacture of rotary endodontic instruments. *Int Endod J* 2005;38:795–801.
33. Chang YZ, Liu MC, Pai CA, Lin CL, Yen KI. Application of non-destructive impedance-based monitoring technique for cyclic fatigue evaluation of endodontic nickel-titanium rotary instruments. *Med Eng Phys* 2011;33:604–9.
34. Lopes HP, Elias CN, Vieira VT, et al. Effects of electropolishing surface treatment on the cyclic fatigue resistance of BioRace nickel-titanium rotary instruments. *J Endod* 2010;36:1653–7.
35. Viana AC, Pereira ES, Bahia MG, Buono VT. The influence of simulated clinical use on the flexibility of rotary ProTaper Universal, K3 and EndoSequence nickel-titanium instruments. *Int Endod J* 2013;46:855–62.

## Figure legends

**Figure 1:** A scheme of the device used for cyclic bending fatigue testing of investigated files: the testing device with its main dimensions, as reported in table 1. RA = radius of the osculating circle for the outer canal wall; RB = radius of the osculating circle for the inner canal wall; alpha = contact angle; a = initial straight portion of the canal; b = curved portion of the canal; c = final straight portion of the canal; p = canal depth; CA = center of the osculating circle for the outer canal wall; CB = center of the osculating circle for the inner canal wall; X,Y = reference system

**Figure 2:** Finite element model (FEM) of PTN X1 file.

**Figure 3:** Comparison between finite element analysis (FEA) and experimental failure locations: vertical red lines denote 95% confidence intervals for the experimental failure locations, highlighted green points denote the location of the element with the smallest number of cycles to failure (NCF).

## Table legends

**Table 1:** Testing configurations adopted for each investigated file.

**Table 2:** FEA = finite element analysis; NCF = number of cycles to failure (mean  $\pm$  SD); \*Length and location from the file tip in mm (mean  $\pm$  SD).

## Multimedia content

**virtual device.mov:** fatigue test simulation by FEA (rigid element representing the testing device in red).

**X1\_stress.mov:** distribution of stress along ProTaper Next X1.

**X2\_stress.mov:** distribution of stress along ProTaper Next X2.

**X3\_stress.mov:** distribution of stress along ProTaper Next X3.

**Table 1**

	<b>Dimensions</b>	<b>X1</b>	<b>X2, X3</b>
1			
2	$R_A [mm]$	5	5
3			
4	$R_B [mm]$	4.6	4.6
5			
6	$alpha [^\circ]$	78	78
7			
8	$a [mm]$	6	6
9			
10	$b [mm]$	7	7
11			
12	$c [mm]$	1	1
13			
14	$p [mm]$	0,6	0,8
15			
16	$X_{CB} [mm]$	+ 0,18	+ 0,04
17			
18	$Y_{CB} [mm]$	- 0,04	- 0,45
19			
20			
21			
22			
23			
24			
25			
26			
27			
28			
29			
30			
31			
32			
33			
34			
35			
36			
37			
38			
39			
40			
41			
42			
43			
44			
45			
46			
47			
48			
49			
50			
51			
52			
53			
54			
55			
56			
57			
58			
59			
60			
61			
62			
63			
64			
65			

Table 2


1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60  
61  
62  
63  
64  
65

Instrument	Cyclic fatigue test		FEA		<i>P</i>	
	NCF	Fracture length*	NCF	Fracture length	NCF	Fracture length
X1	671 ± 170 [592;751]	5.59 ± 0.38 [5.41;5.79]	605	5.77	0.10	0.15
X2	522 ± 67.2 [491;553]	5.46 ± 0.50 [5.23;5.69]	518	5.46	0.89	0.99
X3	334 ± 61.1 [305;362]	5.44 ± 0.16 [5.37;5.51]	329	5.40	0.78	0.48


**Tab. 2** FEA= finite element analysis; NCF = number of cycles to failure (mean ± SD; in square brackets, 95% confidence interval for the mean); \*Length and location from the file tip in mm (mean ± SD).

Figure 1

[Click here to download high resolution image](#)



1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49




**Figure 2**  
[Click here to download high resolution image](#)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49

Figure 3

[Click here to download high resolution image](#)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49



**Video Still**  
[Click here to download high resolution image](#)


- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20
- 21
- 22
- 23
- 24
- 25
- 26
- 27
- 28
- 29
- 30
- 31
- 32
- 33
- 34
- 35
- 36
- 37
- 38
- 39
- 40
- 41
- 42
- 43
- 44
- 45
- 46
- 47
- 48
- 49



**Video Still**

[Click here to download high resolution image](#)


1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49



**Video Still**

[Click here to download high resolution image](#)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49



# Video Still

[Click here to download high resolution image](#)

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49

