

Abstract

121. Plasmonic Nanostructures Integrated in Microfluidic Chips for the Sensitive SERS Detection of miRNAs

Alessandro Chiado' ^{1,*}, Chiara Novara ¹, Niccolò Paccotti ¹, Paola Rivolo ¹, Glenda Biasotto ², Roberto Nisticò ³, Gerolama Condorelli ^{4,5}, Vittorio de Franciscis ⁴, Francesco Geobaldo ¹ and Fabrizio Giorgis ¹

¹ Department of Applied Science and Technology, Politecnico di Torino, Torino, Italy

² LIEC, Institute of Chemistry, São Paulo State University-UNESP, Araraquara, Brazil

³ Department of Applied Science and Technology, Politecnico di Torino, Turin, Italy

⁴ Institute of Endocrinology and Experimental Oncology of Italian National Research Council, Naples, Italy

⁵ Department of Molecular Medicine and Medical Biotechnology, "Federico II" University of Naples, Naples, Italy

* Correspondence: alessandro.chiado@polito.it

Surface-enhanced Raman Scattering (SERS) is a leading technique for the development of innovative biosensors. Indeed, it can be used for the highly sensitive detection of biomolecules in complex biological samples, such as cell extracts or biofluids, taking advantage of plasmonic nanostructures. In this work, an overview of several SERS substrates based on different plasmonic nanostructures such as Ag-PDMS substrates, silvered porous silicon (pSi) on polydimethylsiloxane (PDMS) membranes, 3-D graphene-Ag based aerogels and flower-like Ag nanostructures is presented: each of these platforms have been used for biosensing applications. As a representative example, an elastomeric microfluidic multichamber chip, integrating porous silicon (pSi) membranes decorated with Ag nanoparticles, is here focused on the multiplex detection of miRNAs, short regulatory sequences involved in several diseases, including cancer. In order to exploit the SERS sensitivity, an innovative label-free protocol was developed. The effectiveness of the developed SERS sensing platform was demonstrated by detecting miR-222 into lung cancer cellular extracts, avoiding complex and expensive sample pre-treatments.



© 2019 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).