



POLITECNICO DI TORINO  
Repository ISTITUZIONALE

Controlling bacterial adhesion to titanium surfaces: new strategies for surfaces with tailorable antiadhesive/

*Original*

Controlling bacterial adhesion to titanium surfaces: new strategies for surfaces with tailorable antiadhesive/ antibacterial properties / FERRARIS, SARA; SPRIANO, SILVIA; PERERO, SERGIO; Costa, Piero; Gautier di Confiengo, Giovanna; Cochis, Andrea; Warchomicka, Fernando; Yamaguchi, Seiji; Rimondini, Lia; VERNE', ENRICA; FERRARIS, Monica. - (2020). ((Intervento presentato al convegno 11th World Biomaterials Congress WBC 2020.

*Availability:*

This version is available at: 11583/2828594 since: 2020-05-22T12:54:01Z

*Publisher:*

WBC2020 Conference Secretariat

*Published*

DOI:

*Terms of use:*

openAccess

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

(Article begins on next page)

# Free Session

## *Surface modifications*

WBC2020-2862

### **Controlling bacterial adhesion to titanium surfaces: new strategies for surfaces with tailorable anti-adhesive/antibacterial properties**

Sara Ferraris<sup>1</sup>, Silvia Spriano\*, Sergio Perero<sup>1</sup>, Piero Costa<sup>2</sup>, Giovanna Gautier di Confienigo<sup>3</sup>, Andrea Cochis<sup>4</sup>, Fernando Warchomicka<sup>5</sup>, Seiji Yamaguchi, Lia Rimondini<sup>4</sup>, Enrica Vernè<sup>1</sup>, Monica Ferraris<sup>1</sup>

<sup>1</sup>DISAT, POLITECNICO DI TORINO, <sup>2</sup>INTRAUMA srl, <sup>3</sup>CNR-IMAMOTER, Torino, <sup>4</sup>Università del Piemonte Orientale, Novara, Italy, <sup>5</sup>Graz University of Technology, Graz, Austria

**Please select your preferred method of presentation:** Free Sesion - Oral

**Introduction:** Bacterial contamination is a serious problem in different fields (from everyday life to aerospace) and it becomes critical in biomedical implants: because of increasing bacterial resistance to antibiotics, the development of alternative antibacterial solutions is a high medical need. Implantable biomaterials are frequently associated with infection; a limited colonization and, eventually, a localized active antibacterial action are a challenge for reducing the risk of implant associated infections.

The present research summarizes different strategies aimed at reduction of bacterial adhesion to titanium surfaces moving from simply anti-adhesive surfaces to anti-bacterial and even to bioactive and antibacterial ones.

**Experimental methods:** Titanium and its alloys (mainly Ti6Al4V) were used as substrates.

Different surface treatments were considered in order to improve the biological response of metallic surfaces in specific applications and simultaneously reduce bacterial adhesion. They are briefly summarized in the following Table

- Electron beam structuring has been investigated for the realization of aligned micro-grooves aimed at fibroblast alignment for soft tissue healing. The process induces peculiar surface microstructure and crystallographic structure which can affect bacterial attachment [1]
- Sputtering coating: silica, alumina and zirconia coatings containing metallic silver nanoclusters were deposited on titanium surfaces in order to obtain anti-adhesive coatings for temporary fixation devices with tailorable antibacterial activity [2]
- Bioactive nanotextured titanium oxide layers doped with silver ions/nanoparticles were obtained onto titanium surfaces by means of thermo-chemical treatments in order to combine bioactivity and antibacterial activity in bone contact applications [3, 4].


**Image:**

### Controlling bacterial adhesion to titanium surfaces...



#### E-Beam structuring

- Grooves for fibroblast orientation
- Peculiar microstructure: anti-adhesive for bacteria



#### Sputtering coating

- Silver nanoclusters with tailorable antibacterial activity
- Ceramic matrix can be optimized for reducing tissue adhesion



#### Chemical treatment

- Nanotextured oxide layer with bone bonding ability
- Reduced bacterial adhesion
- Silver ions/nanoparticles with antibacterial activity

**Results and discussions:** As first, all the produced surfaces have mean surface roughness lower than  $0.2 \mu\text{m}$ , reported in literature as threshold to avoid an increase in bacterial adhesion [5] compared to the mirror polished uncoated controls. Moreover, despite of absence of an active antibacterial agent, both EB-structured and nano-textured surfaces are able to reduce bacterial adhesion due to their peculiar microstructure and nanotexture [1, 6]. This strategy is of particular interest because poorly explored: it can be challenging for surfaces with limited bacterial contamination without the introduction of any antibacterial compound. In presence of a high number of bacteria, the anti-adhesive activity is mainly concentrated in the first days, due to the absence of an active bactericidal element. The introduction of silver allows to induce a tailorable antibacterial activity (in fact the amount of Ag can be modulated both in sputtering coating and in thermo-chemical treatments) with limited risk of resistance development. The most challenging point for Ag containing surfaces is a proper modulation of Ag content in order to have simultaneously high antibacterial activity and biocompatibility. The antibacterial activity of these surfaces is based mainly on a ion release mechanism and its duration is strictly related to duration of ion release.

**Conclusions:** Surface modification of titanium surfaces is a versatile strategy in order to obtain anti-adhesive or antibacterial surfaces of interest in the biomedical field. Different processes can be considered depending on the specific application; the control of bacterial adhesion can be coupled with modulation of the interaction between the metallic surface and the biological environment.

**References/Acknowledgements:** [1] S. Ferraris et al. J Mater Proces Tech. 266 (2019) 518–529  
[2] M. Ferraris et al. Appl Surf Sci 396 (2017) 1546-1555  
[3] S. Ferraris et al. Surf Coat Tech 344 (2018) 177–189  
[4] Y. Okuzu Acta Biomaterialia 63 (2017) 383–392  
[5] A. Al-Ahmad et al Arch. Oral Biol. 58 (2013) 1139–1147

[6] S. Ferraris et al. Front. Bioeng. Biotechnol. 7 (2019)103

**Acknowledgments**

MANUNET-2017 and MAECI are acknowledged for founding Easy-Fix and GLOBAL projects

**Disclosure of Interest:** None Declared

**Keywords:** Antibacterial, Coatings, Metallic biomaterials/implants