

Experimental Imaging Issues of a 3-D Microwave Brain Scanner

Original

Experimental Imaging Issues of a 3-D Microwave Brain Scanner / Tobon Vasquez, J.A., Rodriguez-Duarte, D.O., Scapaticci, R., Turvani, G., Bellizzi, G., Joachimowicz, N., Duchêne, B., Casu, M.R., Crocco, L., Vipiana, F.. - ELETTRONICO. - (2021). (ICEAA-IEEE APWC-USNC URSI RSM 2021 Honolulu, HI, USA 9-13 Aug. 2021) [10.1109/ICEAA52647.2021.9539824].

Availability:

This version is available at: 11583/2920756 since: 2021-11-25T12:00:06Z

Publisher:

IEEE

Published

DOI:10.1109/ICEAA52647.2021.9539824

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

IEEE postprint/Author's Accepted Manuscript

©2021 IEEE. Personal use of this material is permitted. Permission from IEEE must be obtained for all other uses, in any current or future media, including reprinting/republishing this material for advertising or promotional purposes, creating new collecting works, for resale or lists, or reuse of any copyrighted component of this work in other works.

(Article begins on next page)



Experimental Imaging Issues of a 3-D Microwave Brain Scanner

J. A. Tobon Vasquez⁽¹⁾, D. O. Rodriguez-Duarte^{*(1)}, R. Scapatucci⁽²⁾, G. Turvani⁽¹⁾, G. Bellizzi⁽³⁾, N. Joachimowicz⁽⁴⁾, B. Duchêne⁽⁵⁾, M. R. Casu⁽¹⁾, L. Crocco⁽²⁾, and F. Vipiana⁽¹⁾.

(1) Politecnico di Torino, Torino, Italy, e-mail: francesca.vipiana@polito.it

(2) IREA-CNR, Napoli, Italy, e-mail: crocco.l@irea.cnr.it

(3) University of Naples Federico II, Napoli, Italy, e-mail: gbellizz@unina.it

(4) Sorbonne Université, CNRS, GeePs, 75252, Paris, France; e-mail: joachimowicz@paris7.jussieu.fr

(5) Université Paris-Saclay, CentraleSupélec, L2S, 91190, Gif-sur-Yvette, France, e-mail: bernard.duchene@l2s.centralesupelec.fr

Microwave imaging is an imaging technique that can be applied to human brain tissues with the potentiality to improve diagnostic ability and to enable earlier diagnosis in the case of trauma and stroke patients [1]. The current most used medical imaging modalities are magnetic resonance imaging (MRI) and computerized tomography (CT) that are assessed and reliable diagnostic tools, however they can be time consuming, costly, not available for bedside monitoring, and harmful due to ionizing radiations (in the CT case only). Hence, microwave imaging can be considered as a complementary diagnostic imaging technique, in particular for bedside and ambulance human head monitoring (see e.g. the systems described in [2], [3]).

In this contribution we present the experimental validation and testing of a 3-D microwave imaging system for brain stroke and hematoma monitoring. The realized system is reported in [4] and is the extension to 3-D imaging of the 2-D system described in [5]. In order to generate 3-D brain images, the antennas/sensors array is placed conformal to upper part of the human head on a helmet-like support. The number, position and orientation of the antennas as well as the choice of the working frequency band and of the dielectric characteristics of the coupling medium have been done accordingly to the rigorous design procedure as described in [6]. The experimental validation and testing of the realized 3-D microwave imaging system is performed with anthropomorphic head phantoms [7] realized with 3-D printed technologies and filled with proper liquid mixtures to mimic the dielectric behavior of the different brain tissues and of the blood area, representing a stroke or a hematoma inside the brain.

This work was supported by the Italian Ministry of University and Research under the PRIN project “MiBraScan”, and by the European Union’s Horizon 2020 Research and Innovation Program under the EMERALD project, Marie Skłodowska-Curie grant agreement No. 764479.

1. A. Fhager, S. Candefjord, M. Elam, and M. Persson, “Microwave diagnostics ahead: Saving time and the lives of trauma and stroke patients,” *IEEE Microwave Mag.*, vol. 19, no. 3, pp. 78–90, May 2018.

2. M. Hopfer, R. Planas, A. Hamidipour, T. Henriksson, and S. Semenov, “Electromagnetic tomography for detection, differentiation, and monitoring of brain stroke: A virtual data and human head phantom study.” *IEEE Antennas Propag. Mag.*, vol. 59, no. 5, pp. 86–97, Oct. 2017.

3. A. T. Mobashsher and A. M. Abbosh, “On-site rapid diagnosis of intracranial hematoma using portable multi-slice microwave imaging system,” *Scientific Reports*, vol. 6, pp. 1–17, Nov. 2016.

4. J. A. Tobon Vasquez, R. Scapatucci, G. Turvani, G. Bellizzi, D. O. Rodriguez-Duarte, N. Joachimowicz, B. Duchene, E. Tedeschi, M. R. Casu, L. Crocco, and F. Vipiana, “A prototype microwave system for 3D brain stroke imaging,” *SENSORS*, vol. 20, May 2020.

5. J. A. Tobon Vasquez, R. Scapatucci, G. Turvani, G. Bellizzi, N. Joachimowicz, B. Duchêne, E. Tedeschi, M. R. Casu, L. Crocco, and F. Vipiana, “Design and experimental assessment of a 2D microwave imaging system for brain stroke monitoring,” *Int. J. Antennas Propag.*, no. Article ID 8065036, p. 12 pages, 2019.

6. R. Scapatucci, J. A. Tobon Vasquez, G. Bellizzi, F. Vipiana, and L. Crocco, “Design and numerical characterization of a low-complexity microwave device for brain stroke monitoring,” *IEEE Trans. Antennas Propag.*, vol. 66, pp. 7328–7338, Dec. 2018.

7. N. Joachimowicz, B. Duchene, C. Conessa, and O. Meyer, “Anthropomorphic breast and head phantoms for microwave imaging,” *Diagnostics*, vol. 85, no. 8, pp. 1–12, Dec. 2018.