

Design and analysis of WELiBot for assisting upper limb motion

Original

Design and analysis of WELiBot for assisting upper limb motion / Jiang, M., Morita, R., Wang, Y., Shinoda, T., Meng, Q., Botta, A., Quaglia, G., Ceccarelli, M., Takeda, Y.. - In: ROBOTICA. - ISSN 0263-5747. - ELETTRONICO. - 43:11(2025), pp. 3921-3941. [10.1017/s0263574725102476]

Availability:

This version is available at: 11583/3003387 since: 2025-09-26T10:06:40Z

Publisher:

Cambridge University Press

Published

DOI:10.1017/s0263574725102476

Terms of use:





This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

RESEARCH ARTICLE

Design and analysis of WELiBot for assisting upper limb motion

Ming Jiang¹ , Ryohei Morita¹, Yang Wang¹, Takashi Shinoda¹, Qizhi Meng¹, Andrea Botta² , Giuseppe Quaglia², Marco Ceccarelli³  and Yukio Takeda¹ 

¹Department of Mechanical Engineering, Institute of Science Tokyo, Tokyo, Japan

²Department of Mechanical and Aerospace Engineering, Politecnico di Torino, Torino, Italy

³Department of Industrial Engineering, University of Rome Tor Vergata, Roma, Italy

Corresponding author: Ming Jiang; Email: jiang.m.889e@m.isct.ac.jp

Received: 25 February 2025; **Revised:** 8 August 2025; **Accepted:** 19 August 2025; **First published online:** 17 September 2025

Keywords: upper limb assistance robot; parallel mechanism; assistive technology; mechanical design; end-effector-type assistance robot

Abstract

Upper limb motor dysfunction significantly impacts daily activities and quality of life for individuals with stroke. Existing assistive robots often struggle to balance portability, ease of use, and motion assistance. This research presents WELiBot, a novel wearable end-effector-type upper limb assistive robot, designed with a 4R-5R parallel mechanism and an arc-shaped guide rail to provide controlled assistance in lifting, reaching, and circumferential motions. The study introduces the conception of the robot design, focusing on its functional requirements and mechanical structure. The kinematic and static characteristics of WELiBot were analyzed to evaluate its feasibility and effectiveness. Based on this design, a prototype with a 1/4 arc-shaped guide rail was fabricated to test motion feasibility and assistance effects. To assess its performance, electromyography experiments were conducted with four healthy participants. The results showed a significant reduction in biceps brachii muscle activity, confirming the robot's ability to reduce user effort. Future work will focus on attaching the guide rail to the body for improved usability and refining the control strategy to enhance motion assistance and adaptability in daily life support applications.

1. Introduction

Each year, more than 12 million people worldwide suffer from strokes, making stroke the third leading cause of death and disability [1, 2]. The sequelae of stroke encompass a wide range of impairments, including motor dysfunction, sensory disturbances, visual field defects, swallowing disorders, higher brain function disorders, and emotional disturbances [3]. Among these, upper limb motor dysfunction is particularly common, often manifesting as muscle weakness or contracture [4], which hinders patients' independence in activities of daily living (ADLs), such as reaching, picking up objects, and holding items [5]. This impairment leads to a decreased quality of life (QOL).

Rehabilitation of the upper limb after a stroke is particularly challenging due to the complexity and variety of movements required for upper limb function. Various neurorehabilitation strategies such as constraint-induced movement therapy (CIMT) [6, 7], mirror therapy [8], and exercise therapy are commonly used to improve motor recovery in stroke patients. It is generally believed that upper limb function recovery occurs primarily within the first few weeks to months after stroke onset [9–11]. Reports indicate that only 11.6% of patients achieve complete functional recovery [11], while 30% to 66% of patients experience persistent upper limb motor dysfunction [9, 12, 13].




Figure 1. Robot types for assisting upper limb motions. (a) Fixed exoskeleton type; (b) Wearable exoskeleton type; (c) Fixed end-effector type; (d) Wearable end-effector type.

In recent years, technology-supported training methods, including assistive and rehabilitation robots, have shown promise in improving upper limb motor function recovery in chronic stroke patients with long-term deficits [14, 15]. These robots also offer the significant advantage of reducing the burden on caregivers [16]. Actively engaging the paretic upper limb in ADL through robotic assistance can help prevent learned non-use, disuse, imbalance between body sides, and poor circulation.

Assistive and rehabilitation robots for upper limbs share similar hardware configurations and can be roughly categorized into three types based on hardware design [17, 18]: fixed exoskeletons (Figure 1a), fixed end-effector systems (Figure 1b), and wearable exoskeletons (Figure 1c). The exoskeleton type is designed to mimic the joint structure of a human upper limb and is typically worn from the shoulder to the forearm. In contrast, the end-effector type is only attached to the forearm. Both fixed exoskeletons and fixed end-effectors are anchored to an external structure, whereas wearable exoskeletons are directly attached to the user's body. Previous research on fixed exoskeleton-type upper limb assistive robots includes chair-mounted [19], and wheelchair-mounted assistive robots [20–22]. These robots vary in their degrees of freedom (DoF), ranging from 3 DoF to 7 DoF, allowing for different levels of joint assistance. Fixed end-effector robots are commonly used in rehabilitation due to their simpler mechanical impedance control, with notable examples such as the MIT-Manus [23], ReoGo™ system [24], and CoCoroe AR2® [25]. Previous studies on wearable exoskeleton-type upper limb assistive robots include the work of Sui et al. [26] and Christensen et al. [27]. Sui et al. [26] developed a robot with 2 DoF at the shoulder, 1 DoF at the elbow, 1 DoF at the forearm, and an additional passive joint with gravity compensation at the sternoclavicular joint. Christensen et al. [27] designed an upper limb assistive robot with 2 DoF at the shoulder and 1 DoF at the elbow, along with passive joints that provide 1 DoF each at the shoulder and scapula. Furthermore, Christensen's robot includes a joint in the spine area that generates restorative force using rubber discs.

These robot types have different advantages and limitations regarding portability, user burden, ease of attachment/detachment, and joint assistance performance. Fixed systems, anchored to external structures, offer low portability but reduce user burden by not placing the robot's weight on the user. In contrast, wearable systems, which are directly attached to the user, are more portable but increase user burden, necessitating lighter designs to mitigate this effect. The distinction between exoskeleton and end-effector systems primarily impacts the ease of attachment and joint assistance. Exoskeletons require precise alignment between the human and robot joint axes, which can cause discomfort if misaligned, complicating attachment. While [28] proposed a method for attachment without adjustment, the increased weight of the robot reduces comfort. End-effector systems, however, avoid these alignment issues, making them easier to attach and detach. Although exoskeletons offer superior joint assistance by providing targeted torque at each joint, end-effector systems, with fewer actuators and less weight, are better suited for minimizing user load.

Among the four possible combinations arising from the above classification, the wearable end-effector type (Figure 1d) has not been extensively studied. Examples of research classified as wearable end-effectors include Honda Motor Co.'s body-weight support walking assist robot [29] and the finger glove developed by Bouzit et al. [30]. However, there appears to be no existing research on wearable end-effector-type robots specifically designed for upper limb assistance or rehabilitation.




Figure 2. Three scenarios in ADLs. (a) Opening a bottle cap; (b) Lifting a box; (c) Cutting vegetables.

To effectively integrate the paretic upper limb into ADLs using an assistive robot, the ability to use the robot anytime and anywhere is crucial for increasing participation frequency. Since no single robot type is excellent in all aspects, the wearable end-effector type, which combines portability features and ease of attachment/detachment, is promising for supporting the paretic upper limb in ADL tasks.

Therefore, this research proposes and designs a wearable end-effector-type upper limb assistive robot, “WELiBot,” to support patients with upper limb hemiplegia after a stroke. In our previous work, we proposed the concept of a parallel mechanism [31] and introduced the robot’s configuration along with a preliminary test [32]. The focus of this paper is to present the design and analysis of “WELiBot” from the conception of the mechanism to the investigation of its effectiveness in assisting the upper limb motion through subjects’ testing.

2. Design conception of the WELiBot

2.1. Target scenarios in ADLs

The design of WELiBot is conceptualized based on three target scenarios in ADLs. Since the use of both hands is challenging for individuals with hemiplegia after a stroke, WELiBot is designed to assist the hemiplegic upper limb. The following three specific scenarios were considered in its conception.

- Opening a bottle cap (Figure 2a): The hemiplegic limb stabilizes the bottle in a fixed position while the non-hemiplegic limb opens the cap.
- Lifting a box (Figure 2b): The hemiplegic limb receives upward assistance, supporting the lifting force while the box is moved upward.
- Cutting vegetables (Figure 2c): The hemiplegic limb is provided with downward assistance to hold the vegetables against the cutting board, while the non-hemiplegic limb operates the knife.

In these scenarios, the hemiplegic upper limb primarily functions as a stabilizing or supportive aid, assisting in fixing or steadying objects. Tasks such as twisting a cap or holding a knife to cut vegetables require precise coordination across multiple joints and controlled force application. Providing robotic assistance for these complex movements would necessitate multiple actuators and sensors, leading to increased weight and cost. Given these constraints, it is more practical and efficient for the non-hemiplegic upper limb to perform these complex tasks, while WELiBot supports the hemiplegic limb in a stabilizing role.

Based on the above considerations, the basic functional requirements for WELiBot are defined. First, it should provide vertical force assistance to support the hemiplegic limb. Second, it should constrain the motion of the hemiplegic limb within a predefined movement range to ensure stability and usability.

2.2. Constrained conditions

In rehabilitation therapy, two primary movements are distinguished in the evaluation and treatment process: the “reaching movement,” where the hand extends forward, and the “lifting movement,” where the




Figure 3. Proposed mechanism of WELiBot.

hand is raised [33]. In addition, movement around the trunk is important for adapting to different environments. Thus, in this study, the mechanism must be capable of producing three types of one-dimensional movements: reaching movement, lifting movement, and circumferential movement.

Previous studies on robotic systems that constrain the end-effector within a specific geometry include the work by Hou et al. [34] and Zhang et al. [35] on upper limb assistance and rehabilitation using a “virtual tunnel.” This approach employs impedance control to create a guided movement sensation, where the user’s hand follows a predefined path. The system minimizes resistance along the set path while increasing resistance perpendicular to the path to constrain motion. Higuchi et al [36] developed the PAS-Arm, a human-compatible assistive arm for lifting tasks, which mechanically constrains the end-effector within a specific geometry. By using a continuously variable transmission, the robot’s joints synchronize to create a guiding plane. These robots rely on multi-joint serial mechanisms, where constraints are achieved by synchronizing joint control or mechanical design.

In this study, we propose that if a parallel mechanism can be partially decoupled with the three one-dimensional movements, a geometric constraint could be achieved by fixing one or two active joints without consuming energy. This approach eliminates the need for synchronized joint control, which is required in serial mechanisms, thereby simplifying the system’s control complexity.

2.3. Proposed mechanism

A 4R-5R parallel mechanism, shown in Figure 3, is proposed to fulfill the functional requirements under the constraints described above. An arc-shaped guide rail is attached around the human trunk, and the end-effector is connected to the hemiplegic forearm through a passive rotational arm cuff. Two sliders, Slider 1 and Slider 2, move along the arc-shaped guide rail, while the end-effector is positioned at the ends of the two chains originating from each slider.

The two chains connecting Slider 1 and Slider 2 are designated as Limb 1 and Limb 2, respectively. Since the sliders move along the arc-shaped guide rail, their motion is kinematically equivalent to a rotational joint at the center of the guide rail. The entire mechanism is represented as $\bar{R}(\hat{R}\hat{R}\hat{R}) - \tilde{R}\tilde{R}(\hat{R}\hat{R}\hat{R})$, where the same superscript symbols indicate parallel joints, and underscored symbols denote active joints. Slider 1, Slider 2, and the rotational joint A_1 serve as active joints.




Figure 4. Coordinates and denotations for the joints and links of the mechanism.

3. Kinematic analysis of the mechanism

3.1. Mobility analysis

Figure 4 illustrate the coordinates and denotations for the joints and links of the mechanism. s_{ij} represents the axis vector, while \hat{S}_{ij} represents the screw, where the first subscript denotes the joint number and the second subscript denotes the limb number. The x -axis is aligned with the direction of OA_1 , the z -axis is aligned with the direction of $s_{1,1} = s_{1,2}$, and the y -axis is determined based on the right-hand coordinate system. The movements of Slider 1 and Slider 2 along the guide rail are kinematically equivalent to rotational movements around the z -axis. The joints exhibit the following parallel relationship that $s_{1,1} \parallel s_{1,2} \parallel s_{2,2}$, $s_{2,1} \parallel s_{3,1} \parallel s_{4,1}$, and $s_{3,2} \parallel s_{4,2} \parallel s_{5,2}$. The angle between the x -axis and OA_2 is defined as ϕ . The angle between OA_2 and A_2C_2 is defined as ψ .

The velocity at point P on Limb 1 and Limb 2 can be expressed as:

$$\hat{S}_{P,1} = \dot{\theta}_{1,1} \hat{S}_{1,1} + \dot{\theta}_{2,1} \hat{S}_{2,1} + \dot{\theta}_{3,1} \hat{S}_{3,1} + \dot{\theta}_{4,1} \hat{S}_{4,1} \tag{1}$$

$$\hat{S}_{P,2} = \dot{\theta}_{1,2} \hat{S}_{1,2} + \dot{\theta}_{2,2} \hat{S}_{2,2} + \dot{\theta}_{3,2} \hat{S}_{3,2} + \dot{\theta}_{4,2} \hat{S}_{4,2} + \dot{\theta}_{5,2} \hat{S}_{5,2} \tag{2}$$

According to Figure 4, the screw representation for each joint on Limb 1 and Limb 2 is given by:

$$\hat{S}_{1,1} = \begin{bmatrix} s_{1,1} \\ \vec{OP} \times s_{1,1} \end{bmatrix} = [0 \ 0 \ 1 \ y_P - y_O \ - (x_P - x_O) \ 0]^T \tag{3}$$

$$\hat{S}_{2,1} = \begin{bmatrix} s_{2,1} \\ \vec{A_1P} \times s_{2,1} \end{bmatrix} = [0 \ 1 \ 0 \ - (z_P - z_{A_1}) \ 0 \ - (x_P - x_{A_1})]^T \tag{4}$$

$$\hat{S}_{3,1} = \begin{bmatrix} s_{3,1} \\ \vec{B_1P} \times s_{3,1} \end{bmatrix} = [0 \ 1 \ 0 \ - (z_P - z_{B_1}) \ 0 \ - (x_P - x_{B_1})]^T \tag{5}$$

$$\hat{S}_{4,1} = \begin{bmatrix} s_{4,1} \\ \vec{C_1P} \times s_{4,1} \end{bmatrix} = [0 \ 1 \ 0 \ 0 \ 0 \ 0]^T \tag{6}$$

$$\hat{S}_{1,2} = \begin{bmatrix} s_{1,2} \\ \vec{OP} \times s_{1,2} \end{bmatrix} = [0 \ 0 \ 1 \ y_P - y_O \ - (x_P - x_O) \ 0]^T \tag{7}$$

$$\hat{S}_{2,2} = \begin{bmatrix} s_{2,2} \\ \vec{A_2P} \times s_{2,2} \end{bmatrix} = [0 \ 0 \ 1 \ y_P - y_{A_2} \ - (x_P - x_{A_2}) \ 0]^T \tag{8}$$

$$\hat{\mathbf{S}}_{3,2} = \begin{bmatrix} \mathbf{s}_{3,2} \\ \vec{A_2P} \times \mathbf{s}_{3,2} \end{bmatrix} = \begin{bmatrix} \cos(\varphi + \psi) \\ \sin(\varphi + \psi) \\ 0 \\ -(z_P - z_{A_2}) \sin(\varphi + \psi) \\ (z_P - z_{A_2}) \cos(\varphi + \psi) \\ \sin(\varphi + \psi)(x_P - x_{A_2}) - \cos(\varphi + \psi)(y_P - y_{A_2}) \end{bmatrix} \tag{9}$$

$$\hat{\mathbf{S}}_{4,2} = \begin{bmatrix} \mathbf{s}_{4,2} \\ \vec{B_2P} \times \mathbf{s}_{4,2} \end{bmatrix} = \begin{bmatrix} \cos(\varphi + \psi) \\ \sin(\varphi + \psi) \\ 0 \\ -(z_P - z_{B_2}) \sin(\varphi + \psi) \\ (z_P - z_{B_2}) \cos(\varphi + \psi) \\ \sin(\varphi + \psi)(x_P - x_{B_2}) - \cos(\varphi + \psi)(y_P - y_{B_2}) \end{bmatrix} \tag{10}$$

$$\hat{\mathbf{S}}_{5,2} = \begin{bmatrix} \mathbf{s}_{5,2} \\ \vec{C_2P} \times \mathbf{s}_{5,2} \end{bmatrix} = [\cos(\varphi + \psi) \sin(\varphi + \psi) 0 0 0 0]^T \tag{11}$$

The reciprocal screw can be obtained using the following equations:

$$\left. \begin{aligned} \hat{\mathbf{S}}_{1,1}^r \circ \hat{\mathbf{S}}_{1,1} &= 0 \\ \hat{\mathbf{S}}_{1,1}^r \circ \hat{\mathbf{S}}_{2,1} &= 0 \\ \hat{\mathbf{S}}_{1,1}^r \circ \hat{\mathbf{S}}_{3,1} &= 0 \\ \hat{\mathbf{S}}_{1,1}^r \circ \hat{\mathbf{S}}_{4,1} &= 0 \end{aligned} \right\} \tag{12}$$

$$\left. \begin{aligned} \hat{\mathbf{S}}_{2,1}^r \circ \hat{\mathbf{S}}_{1,1} &= 0 \\ \hat{\mathbf{S}}_{2,1}^r \circ \hat{\mathbf{S}}_{2,1} &= 0 \\ \hat{\mathbf{S}}_{2,1}^r \circ \hat{\mathbf{S}}_{3,1} &= 0 \\ \hat{\mathbf{S}}_{2,1}^r \circ \hat{\mathbf{S}}_{4,1} &= 0 \end{aligned} \right\} \tag{13}$$

$$\left. \begin{aligned} \hat{\mathbf{S}}_{1,2}^r \circ \hat{\mathbf{S}}_{1,2} &= 0 \\ \hat{\mathbf{S}}_{1,2}^r \circ \hat{\mathbf{S}}_{2,2} &= 0 \\ \hat{\mathbf{S}}_{1,2}^r \circ \hat{\mathbf{S}}_{3,2} &= 0 \\ \hat{\mathbf{S}}_{1,2}^r \circ \hat{\mathbf{S}}_{4,2} &= 0 \\ \hat{\mathbf{S}}_{1,2}^r \circ \hat{\mathbf{S}}_{5,2} &= 0 \end{aligned} \right\} \tag{14}$$

Thus, the reciprocal screw for Limb 1 and Limb 2 is given by:

$$\begin{bmatrix} \hat{\mathbf{S}}_{1,1}^r & \hat{\mathbf{S}}_{2,1}^r & \hat{\mathbf{S}}_{1,2}^r \end{bmatrix} = \begin{bmatrix} 0 & 0 & 0 \\ 0 & 1 & 0 \\ 0 & 0 & 0 \\ 1 & h_1 & -\sin(\varphi + \psi) \\ 0 & 0 & \cos(\varphi + \psi) \\ 0 & x_P - x_O & 0 \end{bmatrix} \tag{15}$$

where h_1 can be assigned an arbitrary value.

The constraint matrix of the mechanism is derived by combining the reciprocal screws of both limbs:

$$\mathbf{J}_C = \begin{bmatrix} \hat{\mathbf{S}}_{1,1}^{rT} \\ \hat{\mathbf{S}}_{2,1}^{rT} \\ \hat{\mathbf{S}}_{1,2}^{rT} \end{bmatrix} = \begin{bmatrix} 0 & 0 & 0 & 1 & 0 & 0 \\ 0 & 1 & 0 & h_1 & 0 & x_P - x_O \\ 0 & 0 & 0 & -\sin(\varphi + \psi) & \cos(\varphi + \psi) & 0 \end{bmatrix} \tag{16}$$




Figure 5. The motion of the mechanism in (a) Vertical direction, (b) Radial direction with a slight change in the height, and (c) Circumferential direction.

By rearranging columns and applying elementary row operations, the constraint matrix can be rewritten as:

$$J_C = \begin{bmatrix} 1 & 0 & 0 & 0 & 0 & 0 \\ 0 & 0 & x_P - x_O & 0 & 1 & 0 \\ 0 & 1 & 0 & 0 & 0 & 0 \end{bmatrix} \tag{17}$$

Since the constraint forces do not perform mechanical work on the mechanism, the motion constraints are obtained as follows:

$$J_C \begin{bmatrix} \mathbf{w} \\ \mathbf{v} \end{bmatrix} = J_C \begin{bmatrix} \omega_x \\ \omega_y \\ \omega_z \\ v_x \\ v_y \\ v_z \end{bmatrix} = 0 \tag{18}$$

$$\left. \begin{aligned} \omega_x &= 0 \\ \omega_z \cdot (x_P - x_O) + v_y &= 0 \\ \omega_y &= 0 \end{aligned} \right\} \tag{19}$$

It is evident that the output link exhibits three DoF. The rotations about the x -axis and y -axis are constrained, while the angular velocity about the z -axis and translational velocity in the y -direction have a proportional relationship with the coefficient $x_P - x_O$.

As shown in Figure 5, modifying the distance between the two sliders causes the end-effector to move approximately in the radial direction with a slight change in the height, corresponding to the “reaching movement.” Rotating the active rotational joint A_1 moves the end-effector in the vertical direction, corresponding to the “lifting movement.” Moving both sliders in the same direction while maintaining their relative distance results in the circumferential motion of the end-effector.

3.2. Displacement analysis

The coordinates, mechanism constants, and variables used for displacement analysis are defined in Figure 6. The O-XYZ coordinate system is a stationary coordinate system with its origin O at the center of the arc-shaped guide rail. The link lengths of Limb 1 are $l_{1,1}, l_{2,1}, l_{3,1}$, while those of Limb 2 are $l_{1,2}, l_{2,2}, l_{3,2}$. The angular displacement of Slider 1 and Slider 2 relative to the X-axis on the XY-plane is given by $\theta_{1,1}$ and $\theta_{2,1}$. The angular displacement of joint A_1 relative to the XY-plane is given by $\theta_{2,1}$.




Figure 6. The coordinates, mechanism constants, and variables for displacement analysis: (a) Mechanism in 3-Dimensions, (b) Projection of the mechanism onto the XY-plane, and (c) Projection of the mechanism onto the uZ- plane.

Figure 6b shows a projection of the mechanism onto the XY-plane, viewed from the positive Z-axis direction. The angle between OC_1 and A_2C_2 is denoted as α . The position of the end-effector P is defined as the intersection of the line passing through C_1 , perpendicular to OC_1 , and the line passing through C_2 , perpendicular to A_2C_2 . The distance between P to A_2C_2 , as well as OC_1 , is presented as l_e . Point Q is the intersection of the extended line of OC_1 and A_2C_2 . A new u -axis is defined along OC_1 in Figure 6b, allowing Limb 1 to be described in the uZ -plane, as shown in Figure 6c. The projection of point P on the u -axis is designated as point R.

For simplicity, interference between the mechanism components, and interference between the mechanism and the human body are neglected in the following calculations.

3.2.1. Forward displacement analysis

In forward displacement analysis, the position of the end-effector $P(X_p, Y_p, Z_p)$ is determined based on the displacements of the active joints $\theta_{1,1}$, $\theta_{2,1}$, and $\theta_{1,2}$.

First, the XY-coordinates of point P are considered. The coordinates of point Q(X_Q, Y_Q) in Figure 6b can be calculated as follows:

$$\left. \begin{aligned} X_Q &= \frac{l_{1,2} \sin \theta_{1,2} - l_{1,2} \cos \theta_{1,2} \tan (\theta_{1,1} - \alpha)}{\tan \theta_{1,1} - \tan (\theta_{1,1} - \alpha)} \\ Y_Q &= \tan \theta_{1,1} \cdot X_Q \end{aligned} \right\} \quad (20)$$

Using X_Q and Y_Q , the XY-coordinates of point P(X_p, Y_p) are expressed as:

$$\left. \begin{aligned} X_p &= X_Q - \frac{l_e}{\sin \frac{\alpha}{2}} \cos \left(\theta_{1,1} - \frac{\alpha}{2} \right) \\ Y_p &= Y_Q - \frac{l_e}{\sin \frac{\alpha}{2}} \sin \left(\theta_{1,1} - \frac{\alpha}{2} \right) \end{aligned} \right\} \quad (21)$$

Next, the Z-coordinate of point P is determined. In Figure 6c, Z_p can be calculated as the intersection between a circle with a radius of $l_{3,1}$ centered at point B_1 and the line $u = u_p$:

$$Z_p = \sqrt{l_{3,1}^2 - (u_p - u_{B,1})^2} + Z_{B,1} \tag{22}$$

where

$$\left. \begin{aligned} u_p &= OR = \sqrt{x_p^2 + y_p^2} \cos \left\{ \tan^{-1} \left(\frac{y_p}{x_p} \right) - \theta_{1,1} \right\} \\ u_{B,1} &= l_{1,1} + l_{2,1} \cos \theta_{2,1} \\ Z_{B,1} &= l_{2,1} \sin \theta_{2,1} \end{aligned} \right\} \tag{23}$$

3.2.2. Inverse displacement analysis

In inverse displacement analysis, the displacements of the active joints $\theta_{1,1}$, $\theta_{2,1}$, and $\theta_{1,2}$ are determined based on the position of the end-effector (X_p, Y_p, Z_p).

The angular displacement $\theta_{1,1}$ and $\theta_{1,2}$ of Slider 1 and Slider 2 can be calculated using X_p and Y_p .

$$\theta_{1,1} = \tan^{-1} \left(\frac{X_p}{Y_p} \right) - \sin^{-1} \left(\frac{l_e}{\sqrt{X_p^2 + Y_p^2}} \right) \tag{24}$$

Using the calculated $\theta_{1,1}$, the coordinates of point Q(X_Q, Y_Q) can be expressed as follows:

$$\left. \begin{aligned} X_Q &= X_p + \frac{l_e}{\sin \frac{\alpha}{2}} \cos \left(\theta_{1,1} - \frac{\alpha}{2} \right) \\ Y_Q &= X_p + \frac{l_e}{\sin \frac{\alpha}{2}} \sin \left(\theta_{1,1} - \frac{\alpha}{2} \right) \end{aligned} \right\} \tag{25}$$

The value of $\theta_{1,2}$ is determined as follows:

$$\theta_{1,2} = \pm \cos^{-1} \left\{ \frac{Y_Q \cos (\theta_{1,1} - \alpha) - X_Q \sin (\theta_{1,1} - \alpha)}{l_{1,2}} \right\} + \left(\theta_{1,1} - \alpha + \frac{\pi}{2} \right) \tag{26}$$

From the sign of the first term on the right-hand side, it is evident that the inverse kinematic solution for $\theta_{1,2}$ has two possible values. The positive solution results in link A_2B_2 being positioned inside the arc-shaped guide rail, which is unsuitable. Therefore, the negative solution is selected for the subsequent analysis. The value of $\theta_{2,1}$ can be calculated as follows, based on Figure 6c.

$$\theta_{2,1} = \angle PA_1R - \angle PA_1B_1 = \tan^{-1} \left(\frac{Z_p}{u_{AR}} \right) - \cos^{-1} \left(\frac{u_{AR}^2 + Z_p^2 + l_{2,1}^2 - l_{3,1}^2}{2l_{2,1}\sqrt{u_{AR}^2 + Z_p^2}} \right) \tag{27}$$

where

$$u_{AR} = A_1R = \sqrt{X_p^2 + Y_p^2} \cos \left\{ \tan^{-1} \left(\frac{Y_p}{X_p} \right) - \theta_{1,1} \right\} - l_{1,1} \tag{28}$$

3.3. Workspace analysis

Similar to the displacement analysis, interference between the mechanism’s joints and links, as well as interference between the mechanism and the human body, are neglected in this study. In addition, the constraint $\theta_{1,1} < \theta_{1,2}$ is applied to the sliders since the two sliders cannot pass each other. In practice, two sliders contact each other at a certain angular difference. For this analysis, however, the sliders are theoretically treated as points.

Since the XY -coordinates (X_p, Y_p) of the end-effector are determined solely by the slider positions $\theta_{1,1}$ and $\theta_{1,2}$, the workspace of the mechanism is analyzed in the XY -plane and Z -coordinate, respectively. Let r_p represent the distance between the origin point O and the end-effector P , with its maximum and minimum values denoted as $r_{p,max}$, $r_{p,min}$, respectively.

This can be analytically derived based on the results of the displacement analysis. In Figure 6b, the shape of the quadrilateral PC_1QC_2 remains constant, regardless of $\theta_{1,1}$ and $\theta_{1,2}$. Therefore, when the distance r_Q between the origin point O and point Q reaches its maximum or minimum value, the distance r_p also takes its respective maximum or minimum value. Moreover, since the mechanism is symmetric about the Z -axis, when evaluating the maximum and minimum values of r_Q , $\theta_{1,1}$ can be fixed at 0, and only $\theta_{1,2}$ is considered as a variable. By substituting $\theta_{1,1} = 0$ into Eq. (20), we obtain the Eq. (29):

$$\left. \begin{aligned} X_Q &= \frac{l_{1,2}}{\sin \alpha} (\cos \alpha \sin \theta_{1,2} + \sin \alpha \cos \theta_{1,2}) = \frac{l_{1,2}}{\sin \alpha} \cdot \sin (\theta_{1,2} + \alpha) \\ Y_Q &= 0 \end{aligned} \right\} \quad (29)$$

so that,

$$r_Q = \sqrt{X_Q^2 + Y_Q^2} = \frac{l_{1,2}}{\sin \alpha} \cdot \sin (\theta_{1,2} + \alpha) \quad (30)$$

Therefore, r_Q reaches its maximum value when the term $\sin(\theta_{1,2} + \alpha)$ on the right-hand side reaches its maximum value, satisfying $OA_2 \perp A_2C_2$:

$$\theta_{1,2} = \frac{\pi}{2} - \alpha \quad (31)$$

Similarly, r_Q reaches its minimum value when $\sin(\theta_{1,2} + \alpha)$ takes its minimum value. Since $\sin(\theta_{1,2} + \alpha)$ increases monotonically within the range $0 \leq \theta_{1,2} \leq \pi/2 - \alpha$, its minimum value occurs when $\theta_{1,2} = 0$. Therefore, $r_p = r_{p,min}$ occurs when $\theta_{1,1} = \theta_{1,2}$.

Based on this, $r_{p,max}$ and $r_{p,min}$ can be determined as follows:

$$\left. \begin{aligned} r_{p,max} &= \sqrt{\left(\frac{l_{1,2}}{\sin \alpha} - \frac{l_e}{\tan (\alpha/2)} \right)^2 + l_e^2} \left(\text{when } \theta_{1,2} = \theta_{1,1} + \frac{\pi}{2} - \alpha \right) \\ r_{p,min} &= \sqrt{\left(l_{1,2} - \frac{l_e}{\tan (\alpha/2)} \right)^2 + l_e^2} \left(\text{when } \theta_{1,1} = \theta_{1,2} \right) \end{aligned} \right\} \quad (32)$$

The workspace in the XY -plane is represented as a doughnut-shaped region, indicated by the gray area in Figure 7a.

Next, the maximum value $Z_{p,max}$ and minimum value $Z_{p,min}$ are determined, given values of X_p and Y_p . First, consider only Limb 1. Figure 7b represents Limb 1 in the uZ -plane. Z_p reaches its maximum value $Z_{p,max,1}$ or minimum value $Z_{p,min,1}$, when points A_1, B_1 , and C_1 are aligned in a straight line. Therefore, $Z_{p,max,1}$ and $Z_{p,min,1}$ can be expressed as follows:

$$\begin{aligned} Z_{p,max,1} &= \sqrt{(l_{2,1} + l_{3,1})^2 - u_{AR}^2} \\ Z_{p,min,1} &= -Z_{p,max,1} = -\sqrt{(l_{2,1} + l_{3,1})^2 - u_{AR}^2} \end{aligned} \quad (33)$$

The value of u_{AR} can be determined using (28).

Similarly, considering only Limb 2, Figure 7c represents Limb 2 in the vZ_2 -plane, where the Z_2 -axis is defined as the line parallel to the Z -axis, passing through point A_2 , and the v -axis corresponds to the A_2C_2 direction as shown in Figure 6b. For Limb 2, Z_p reaches its maximum value $Z_{p,max,2}$ or minimum value $Z_{p,min,2}$ when points A_2, B_2 , and C_2 are aligned in a straight line. Therefore, $Z_{p,max,2}$ and $Z_{p,min,2}$ can




Figure 7. The workspace of the mechanism in (a) XY-plane and (b) Z-axis direction.

be expressed as follows:

$$\left. \begin{aligned} Z_{P,max,2} &= \sqrt{(l_{2,2} + l_{3,2})^2 - v_{AS}^2} \\ Z_{P,min,2} &= -Z_{P,max,2} = -\sqrt{(l_{2,2} + l_{3,2})^2 - v_{AS}^2} \end{aligned} \right\} \quad (34)$$

where

$$v_{AS} = \sqrt{\{x_P - l_e \sin(\theta_{1,1} - \alpha) - l_{1,2} \cos \theta_{1,2}\}^2 + \{y_P + l_e \cos(\theta_{1,1} - \alpha) - l_{1,2} \sin \theta_{1,2}\}^2} \quad (35)$$

Based on this, considering both Limb 1 and Limb 2, $Z_{P,max}$ is constrained by the smaller value between $Z_{P,max,1}$ and $Z_{P,max,2}$, while $Z_{P,min}$ is constrained by the larger value between $Z_{P,min,1}$ and $Z_{P,min,2}$. Therefore, $Z_{P,max}$ and $Z_{P,min}$ can be expressed as follows:

$$\left. \begin{aligned} Z_{P,max} &= \min(Z_{P,max,1}, Z_{P,max,2}) \\ Z_{P,min} &= \max(Z_{P,min,1}, Z_{P,min,2}) \end{aligned} \right\} \quad (36)$$

In addition, the end-effector cannot reach within a circular region centered at A_1 of Limb1 and A_2 of Limb2, with radii $r = |l_{2,1} - l_{3,1}|$ and $r = |l_{2,2} - l_{3,2}|$, respectively. Therefore, these regions must be excluded from the workspace.

Based on the above analysis, the workspace in three-dimensional space can be determined by obtaining the boundary of the workspace from the XY-plane and the Z-axis direction while excluding the unreachable regions. Since the mechanism is symmetric about the Z-axis, the three-dimensional workspace can be obtained by rotating the XZ plane workspace 360° around the Z-axis.

4. Static analysis

The mechanism consists of seven links, excluding the fixed circular guide rail. First, the equations of force and moment equilibrium are established for each of the seven segments, as illustrated in Figure 8.

Let F_i and M_i represent the force and moment acting on each link, respectively. The mass of each link is denoted as m_i , and the center of mass is represented by the position vector G_i . The distance between the origin O and each slider is denoted as r_i , while the length of each link is l_i . g represents the gravitational acceleration. The equilibrium equations of force and moment acting on the links are expressed as follows:

$$\left. \begin{aligned} F_{S1} - F_{A1} + m_{SA,1}g &= 0 \\ M_{S1} - M_{A1} + \{(A_1 - S_1) \times (-F_{A1})\} + \{(G_{SA,1} - S_1) \times m_{SA,1}g\} &= 0 \end{aligned} \right\} \quad (37)$$




Figure 8. Free body diagram of the mechanism.

$$\left. \begin{aligned} F_{A1} - F_{B1} + m_{AB,1}g &= 0 \\ M_{A1} - M_{B1} + \{(B_1 - A_1) \times (-F_{B1})\} + \{(G_{AB,1} - A_1) \times m_{AB,1}g\} &= 0 \end{aligned} \right\} \quad (38)$$

$$\left. \begin{aligned} F_{B1} - F_{C1} + m_{BC,1}g &= 0 \\ M_{B1} - M_{C1} + \{(C_1 - B_1) \times (-F_{C1})\} + \{(G_{BC,1} - B_1) \times m_{BC,1}g\} &= 0 \end{aligned} \right\} \quad (39)$$

$$\left. \begin{aligned} F_{S2} - F_{A2} + m_{SA,2}g &= 0 \\ M_{S2} - M_{A2} + \{(A_2 - S_2) \times (-F_{A2})\} + \{(G_{SA,2} - S_2) \times m_{SA,2}g\} &= 0 \end{aligned} \right\} \quad (40)$$

$$\left. \begin{aligned} F_{A2} - F_{B2} + m_{AB,2}g &= 0 \\ M_{A2} - M_{B2} + \{(B_2 - A_2) \times (-F_{B2})\} + \{(G_{AB,2} - A_2) \times m_{AB,2}g\} &= 0 \end{aligned} \right\} \quad (41)$$

$$\left. \begin{aligned} F_{B2} - F_{C2} + m_{BC,2}g &= 0 \\ M_{B2} - M_{C2} + \{(C_2 - B_2) \times (-F_{C2})\} + \{(G_{BC,2} - B_2) \times m_{BC,2}g\} &= 0 \end{aligned} \right\} \quad (42)$$

$$\left. \begin{aligned} F_{C1} + F_{C2} + F_E + m_Eg &= 0 \\ M_{C1} + M_{C2} + M_E + \{(C_2 - C_1) \times (F_{C2})\} + \{(P - C_1) \times (F_E + m_Eg)\} &= 0 \end{aligned} \right\} \quad (43)$$

In addition to these equilibrium equations, the condition that the moment around the axis of each passive revolute joint ($B_1, C_1, A_2, B_2,$ and C_2) is zero is applied:

$$\left. \begin{aligned} s_{3,1} \cdot M_{B1} &= 0 \\ s_{4,1} \cdot M_{C1} &= 0 \\ s_{2,2} \cdot M_{A2} &= 0 \\ s_{3,2} \cdot M_{A2} &= 0 \\ s_{4,2} \cdot M_{B2} &= 0 \\ s_{5,2} \cdot M_{C2} &= 0 \end{aligned} \right\} \quad (44)$$

Given that the position vectors of the joints and centers of mass, the segment masses, and the external forces and moments acting on the end-effector are known, the number of unknown joint forces and moments matches the number of equations, both being 48. Therefore, Eqs. (37) to (44) are solvable.




Figure 9. The fabricated prototype: (a) Configuration of the prototype (b) Prototype attached to the human arm.

5. Prototype

A prototype was fabricated to evaluate the effectiveness of motion assistance through experiments involving upper limb attachment. Figure 9 shows a photograph of the developed prototype, and Table I lists the mechanism parameters. In this study, the prototype was not intended for actual attachment to the body or practical application. It was for verifying the mechanical feasibility and baseline assistive performance. Therefore, the circular guide rail was designed as an arc with a central angle of 90 degrees and was fixed to a stand. Aluminum was used as the primary material for most parts, while carbon fiber-reinforced nylon (Onyx), produced using a 3D printer, was used for the arm cuff and some simple spacers. The weight of the prototype was around 3.9 kg.

Figure 10a shows the structure of the circular guide rail. The radial cross-section has an H-shape with 45-degree inclined surfaces. A flexible rack from Kohara Gear Industry Co., Ltd. is fixed to the outer surface of the guide rail. Figure 10b and 10c show the structure of Slider 1 and Slider 2, respectively. The roller structure is common to both sides, as illustrated in Figure 10e. The rollers are mounted at a 45-degree tilt so that they contact the inclined surfaces of the circular guide rail. Roller tighten screws allow the rollers to be pressed against the guide rail by adjusting the screws in the direction of the blue arrow. Figure 10d shows the end-effector and arm cuff components. The arm cuff includes two passive revolute joints around the axes indicated by the black dashed lines. Rotational motion around the axis shown by the red dashed line is accommodated by slippage between the arm cuff and the forearm or between the forearm and clothing. For simplicity, these rotations were not incorporated into the mechanism as revolute joints. The prototype also included several design features for safety. Mechanical stoppers were mounted at both ends of the arc-shaped guide rail to prevent the sliders from detaching during operation. An emergency stop button was positioned at the end of the guide, allowing immediate interruption of motion if necessary. The arm cuff was equipped with passive rotational joints to reduce the risk of excessive torque on the user's arm, and a Velcro fastening tape was used to enable quick detachment when needed.

Figure 11 shows the circuit configuration of the prototype, and Table II lists the model numbers of the major components. The motor power supply was provided externally using a 12V DC stabilized power source.

Table I. Parameters of the prototype.

Parameters	Values
$l_{1,1}$	0.266 m
$l_{2,1}$	0.250 m
$l_{3,1}$	0.250 m
$l_{1,2}$	0.280 m
$l_{2,2}$	0.280 m
$l_{3,2}$	0.280 m
l_e	0.037 m
α	30 degrees




Figure 10. The configuration of the prototype: (a) arc-shape guide rail (b) Slider 1 (c) Slider 2 (d) Arm cuff (e) Rollers and the tighten screw.




Figure 11. The circuit configuration.

6. Experiments

To evaluate the concept and assistance performance of WELiBot, experiments were conducted with four healthy participants (age: 26 ± 4.5 , weight: 65 ± 8.2 kg, height: 172 ± 7.6 cm) to analyze the differences in muscle activity levels under two conditions: with and without robotic assistance. Each participant

Table II. Main components of the prototype.

Component	Model	Manufacturer
Microcomputer	STM32F446RET6	STMicroelectronics
Motor driver 1	MD20A	Cytron Technologies
Motor driver 2	MD10C	Cytron Technologies
Current sensor	ACS711KEXLT-31AB-T	Allegro MicroSystems
Motor at joint A_1	MS-555VC-5524	Mabuchi Motor
Motor on the sliders	RS-385PH-2465	Mabuchi Motor
Encoder	AMT103-V	CUI Devices




Figure 12. The block diagram of the PD controller.

repeated the specified motion five times, and the muscle activities of the anterior deltoid, biceps brachii, and infraspinatus were measured using electromyography (EMG) sensors (Biometrics, UK).

In addition to EMG analysis, a motion measurement experiment was conducted to assess the displacements of WELiBot quantitatively. Four representative upper limb movements were performed: circumferential, raising, reaching forward, and a combined motion of the first three. Each motion was executed five times without the participants, and the relative displacements Δx , Δy , and Δz of the end-effector (endpoint minus starting point) were recorded using an OptiTrack (NaturalPoint, Inc., USA) motion capture system.

6.1. Experimental setup

In both the assistance experiment and the motion measurement, the arm cuff of the robot was moved along a straight trajectory with PD control, following a trapezoidal acceleration and deceleration profile with an acceleration of $a = 0.5 \text{ m/s}^2$ and a maximum velocity of $v_{\max} = 0.3 \text{ m/s}$. The total motion time was about 1.67 s. Muscle activity measurements and the robot’s motion were synchronized using a trigger switch (Biometrics Ltd, UK). The block diagram of the controller is shown in Figure 12. In the block diagram, x_d is the target trajectory of the end-effector, and θ_d and θ_a indicate the target angle and measured angle of each active joint, respectively. \dot{x}_d and $\dot{\theta}_d$ are the corresponding velocity and angular velocity. V is the driving voltage. F_{assist} is the assistive force. s is the Laplace variable.

In the condition without robot assistance, participants were instructed to move their arms along the same trajectory as in the robot-assisted condition, maintaining the same start and end positions. The motion was guided by a metronome, set at a beat interval of 60/72 s. The specific motion was performed over two beats, ensuring consistency with the robot-assistance condition.

6.2. Results of muscle activity

In the assistance experiment, the specified motion involved a straight-line movement from the position at (0.24, 0.32, 0) to (0.1, 0.37, 0.30) (m) in the coordinate shown in Figure 9a. This motion simulates




Figure 13. RMS from four subjects under the condition with and without robot assistance.

the daily task of picking up an object from a table and holding it in a static position. Figure 9b shows an example of the experimental setup with robot assistance.

The EMG voltage $V_{EMG}(t)$ at each moment was computed from the raw EMG data $e(t)$ using the following formula, with a window length $T = 300$ ms:

$$V_{EMG}(t) = \sqrt{\frac{1}{T} \int_{-T/2}^{T/2} e^2(t + \tau) d\tau} \tag{45}$$

This process was performed for each of the five trials under the same conditions, and the average $V_{EMG}(t)$ across the trials is shown in Figure 13. Moreover, the mean absolute value (MAV) of $V_{EMG}(t)$ was calculated over the entire duration of 1.67 s, representing the overall muscle activity for each trial. An independent two-tailed t -test has been conducted to compare the difference in conditions with and without robot assistance. The results shown in Figure 14 indicate that the muscle activities of the infraspinatus, anterior deltoid, and biceps brachii were reduced in the robot-assistance condition, confirming the effectiveness of assistance. In particular, for all four subjects, the biceps brachii muscle activity was significantly decreased.




Figure 14. Comparison of EMG data from four subjects.

6.3. Results of measured end-effector displacement

The measured coordinates of the starting and endpoint for each motion are summarized in Table III.

The standard deviations of relative displacement along the x, y, and z axes ($\sigma_x, \sigma_y, \sigma_z$) were computed, and a 3D repeatability radius was defined as:

$$R = \sqrt{\sigma_x^2 + \sigma_y^2 + \sigma_z^2} \tag{46}$$

Table IV presents the mean displacement, standard deviations, and repeatability results for the four motions.

The repeatability was 0.43 mm for circumferential motion, 0.66 mm for rising motion, 0.46 mm for reaching motion, and 0.40 mm for combined motion. These results indicate that WELiBot consistently reproduced similar motion displacements across repeated motion tasks, demonstrating stable and reliable actuation performance.

7. Discussion

The experimental results demonstrated that WELiBot effectively reduced muscle activity, particularly in the biceps brachii, confirming its potential for upper limb assistance. The displacement measurement shows that robots can realize reliable actuation performance. Despite its effectiveness, this study has several limitations related to both analysis and experimental setup.

One major limitation of the present study lies in the experimental participants. The intended users of WELiBot are individuals with hemiplegia. However, the current study was conducted with healthy individuals, which does not fully reflect the challenges faced by individuals with upper limb impairments. To develop an assistive device truly suitable for hemiplegic upper limb rehabilitation, WELiBot is expected to improve in the following three directions. First, the current prototype is fixed on a stand and is not wearable. For a user with limb impairment, the robot must prioritize lightweight construction, ease of attachment and removal, and minimal setup burden, ideally allowing the user to wear the system independently or with minimal assistance from a caregiver. Second, an appropriate control strategy for WELiBot is needed. The symptoms and movement patterns of patients with hemiplegia vary significantly across individuals. Therefore, the robot is expected to incorporate adaptive control strategies, such as impedance control, EMG-based control [37], or Assist-As-Needed (AAN) frameworks [38]. These approaches aim to assist each user according to their residual function and promote active use of the affected limb, which is considered critical in enhancing neuroplasticity and functional recovery.

Table III. Measured coordinates of the starting point and the endpoint.

Motion Type	Measured mean starting point (m)	Measured mean endpoint (m)
Circumferential	(0.24, 0.20, 0.23)	(0.18, 0.27, 0.22)
Rising	(0.18, 0.28, 0.24)	(0.18, 0.30, 0.42)
Reaching	(0.24, 0.20, 0.23)	(0.19, 0.28, 0.25)
Combined	(0.21, 0.25, 0.23)	(0.17, 0.29, 0.41)

Table IV. Measured end-effector displacement.

Motion Type	Mean Δx (mm)	Mean Δy (mm)	Mean Δz (mm)	σ_x (mm)	σ_y (mm)	σ_z (mm)	Repeatability (mm)
Circumferential	-63.08	62.35	-2.63	0.15	0.37	0.16	0.43
Rising	-0.12	12.61	182.49	0.25	0.41	0.45	0.66
Reaching	-51.37	80.15	18.81	0.21	0.35	0.22	0.46
Combined	-39.19	35.15	180.27	0.28	0.27	0.08	0.40

Third, the WELiBot design utilized passive rotational joints in the arm cuff to enhance safety for subject experiments. Nevertheless, therapists have pointed out that such passive mechanisms may not provide adequate support for patients with muscle spasticity or joint stiffness, which are common in hemiplegia. During upper limb rehabilitation, therapists typically stabilize the patient's wrist manually to facilitate movement at the shoulder and elbow. Hence, future iterations of WELiBot should include a redesigned wrist interface that provides secure, adjustable, and clinically informed support tailored to hemiparetic conditions. In addition, further studies involving a larger sample size and participants with upper limb impairments are necessary to validate the effectiveness of WELiBot in real-world ADL scenarios.

In terms of kinematic and static analysis, the study did not account for direct interactions between the robot and the human body. Since WELiBot is designed as a wearable device to assist with ADLs in individuals with motor impairments, the current prototype does not fully address practical challenges associated with real-world use. In addition, the workspace analysis was based on geometric constraints, without incorporating human anatomical factors, such as joint range of motion and individual movement variability. These factors may influence the robot's effectiveness and usability and should be integrated into future design considerations.

Regarding the experimental setup, the prototype was not attached to the body, as the guide rail was fixed to a stationary stand, and only one-quarter of the rail was fabricated. This limited the assessment of the robot's full motion range and usability. Attaching the guide rail directly to the human body may introduce additional constraints, dynamic interactions, and comfort-related factors, which require further investigation. In addition, the experiment was limited to a predefined straight-line motion, constrained by the workspace range of the fabricated prototype. Future studies should investigate the WELiBot's performance with a full-scale prototype.

While a fully wearable version of WELiBot remains a long-term objective, alternative configurations such as a frame-mounted or semi-mobile system may be more practical for some users, especially elderly individuals or those with severe hemiplegia, because such users may face difficulties tolerating wearable systems. A fixed-stand configuration, when properly designed, can still provide meaningful motion assistance while improving usability and safety. This design consideration is also worth exploring further in future development stages to balance functionality with accessibility in rehabilitation applications.

8. Conclusion

This study mainly introduced the design of WELiBot, a novel wearable end-effector-type upper limb assistive robot designed to provide effective support for individuals with upper limb motor

impairments. The research began with the conception of robot design, defining its functional requirements and mechanical structure. A 4R-5R parallel mechanism was developed to enable controlled assistance in lifting, reaching, and circumferential motions while ensuring ease of attachment and portability. Kinematic and workspace analyses confirmed that the mechanism provides a well-defined, toroidal workspace centered around the user's torso, meeting motion assistance requirements. In addition, static analysis provided insights into force distribution, aiding in the selection of design parameters. To assess mechanical feasibility, a prototype with a quarter arc-shaped guide rail was fabricated and tested. Muscle activity experiments conducted with four healthy participants demonstrated a significant reduction in biceps brachii muscle activity compared to unassisted movement, confirming WELiBot's effectiveness in reducing physical effort. Future research will focus on attaching the guide rail to the body to improve usability, refining the control strategy for enhanced motion assistance, and conducting further daily task-based experiments to evaluate the robot's real-world effectiveness.

Acknowledgements. This research is an extended version of the conference paper "Design and Preliminary Testing of WELiBot: A Wearable End-Effector Type Upper Limb Assistive Robot," published in *Mechanisms and Machine Science*, doi:[10.1007/978-3-031-64569-3_9](https://doi.org/10.1007/978-3-031-64569-3_9).

Author contributions. Ming Jiang and Ryohei Morita contributed to study methodology, investigation, software development, data curation, visualization, manuscript drafting, review, and editing. Yang Wang and Takashi Shinoda contributed to study methodology, investigation, software development, data curation, and manuscript review and editing. Qizhi Meng, Andrea Botta, Giuseppe Quaglia, and Marco Ceccarelli provided insights into the methodology, mechanical design workspace analysis, and manuscript review and editing. Yukio Takeda supervised the study, oversaw the study methodology, facilitated resource acquisition, and contributed to manuscript review and editing.

Financial support. This work was partly supported by the Mizuho Foundation for the Promotion of Sciences and the Fujikura Foundation.

Competing interests. The authors declare no conflicts of interest exist.

Ethical approval. All experiments in this study were conducted following the Ethical Guidelines for Life Science and Medical Research Involving Human Subjects. The protocol for this research received approval from the Human Subjects Research Ethics Review Committee of the Institute of Science Tokyo, approval number 2024204.

References

- [1] V. L. Feigin, B. A. Stark, C. O. Johnson, G. A. Roth, C. Bisignano, G. G. Abady, M. Abbasifard, M. Abbasi-Kangevari, F. Abd-Allah, V. Abedi, A. Abualhasan, N. M. E. Abu-Rmeileh, A. I. Abushouk, O. M. Adebayo, G. Agarwal, P. Agasthi, B. O. Ahinkorah, S. Ahmad, S. Ahmadi, Y. Ahmed Salih, B. Aji, S. Akbarpour, R. O. Akinyemi, H. Al Hamad, F. Alahdab, S. M. Alif, V. Alipour, S. M. Aljunid, S. Almoustanyir, R. M. Al-Raddadi, R. Al-Shahi Salman, N. Alvis-Guzman, R. Anuceanu, D. Anderlini, J. A. Anderson, A. Ansar, I. C. Antonazzo, J. Arabloo, J. Ärnlov, K. D. Artanti, Z. Aryan, S. Asgari, T. Ashraf, M. Athar, A. Atreya, M. Ausloos, A. A. Baig, O. C. Baltatu, M. Banach, M. A. Barboza, S. L. Barker-Collo, T. W. Bärnighausen, M. T. U. Barone, S. Basu, G. Bazmandegan, E. Beghi, M. Beheshti, Y. Béjot, A. W. Bell, D. A. Bennett, I. M. Bensenor, W. M. Bezabhe, Y. M. Bezabih, A. S. Bhagavathula, P. Bhardwaj, K. Bhattacharyya, A. Bijani, B. Bikbov, M. M. Birhanu, A. Bolor, A. Bonny, M. Brauer, H. Brenner, D. Bryazka, Z. A. Butt, F. L. Caetano dos Santos, I. R. Campos-Nonato, C. Cantu-Brito, J. J. Carrero, C. A. Castañeda-Orjuela, A. L. Catapano, P. A. Chakraborty, J. Charan, S. G. Choudhari, E. K. Chowdhury, D.-T. Chu, S.-C. Chung, D. Colozza, V. M. Costa, S. Costanzo, M. H. Criqui, O. Dadras, B. Dagneu, X. Dai, K. Dalal, A. A. M. Damasceno, E. D'Amico, L. Dandona, R. Dandona, J. Darega Gela, K. Davletov, V. De la Cruz-Góngora, R. Desai, D. Dhamnetiya, S. D. Dharmaratne, M. L. Dhimal, M. Dhimal, D. Diaz, M. Dichgans, K. Dokova, R. Doshi, A. Douiri, B. B. Duncan, S. Eftekharzadeh, M. Ekhloenetale, N. El Nahas, I. Y. Elgendy, M. Elhadi, S. I. El-Jaafary, M. Endres, A. Y. Endries, D. A. Erku, E. J. A. Faraon, U. Farooque, F. Farzadfar, A. H. Feroze, I. Filip, F. Fischer, D. Flood, M. M. Gad, S. Gaidhane, R. Ghanei Gheshlagh, A. Ghashghaee, N. Ghith, G. Ghozali, S. Ghozy, A. Gialluisi, S. Giampaoli, S. A. Gilani, P. S. Gill, E. V. Gnedovskaya, M. Golechha, A. C. Goulart, Y. Guo, R. Gupta, V. B. Gupta, V. K. Gupta, P. Gyanwali, N. Hafezi-Nejad, S. Hamidi, A. Hanif, G. J. Hankey, A. Hargono, A. Hashi, T. S. Hassan, H. Y. Hassen, R. J. Havmoeller, S. I. Hay, K. Hayat, M. I. Hegazy, C. Herteliu, R. Holla, S. Hostiuc, M. Househ, J. Huang, A. Humayun, B.-F. Hwang, L. Iacoviello, I. Iavicoli, S. E. Ibitoye, O. S. Ilesanmi, I. M. Ilic, M. D. Ilic, U. Iqbal, S. S. N. Irvani, S. M. S. Islam, N. E. Ismail, H. Iso, G. Isola, M. Iwagami, L. Jacob, V. Jain, S.-I. Jang, S. K. Jayapal, S. Jayaram, R. Jayawardena, P. Jeemon, R. P. Jha, W. D. Johnson, J. B. Jonas, N. Joseph, J. J. Jozwiak, M. Jürisson, R. Kalani, R. Kalhor,

- Y. Kalkonde, A. Kamath, Z. Kamiab, T. Kanchan, H. Kandel, A. Karch, P. D. M. C. Katoto, G. A. Kayode, P. Keshavarz, Y. S. Khader, E. A. Khan, I. A. Khan, M. Khan, M. A. B. Khan, M. N. Khatib, J. Khubchandani, G. R. Kim, M. S. Kim, Y. J. Kim, A. Kisa, S. Kisa, M. Kivimäki, D. Kolte, A. Koolivand, S. L. Koulmane Laxminarayana, A. Koyanagi, K. Krishan, V. Krishnamoorthy, R. V. Krishnamurthi, G. A. Kumar, D. Kusuma, C. La Vecchia, B. Lacey, H. M. Lak, T. Lallukka, S. Lasrado, P. M. Lavados, M. Leonardi, B. Li, S. Li, H. Lin, R.-T. Lin, X. Liu, W. D. Lo, S. Lorkowski, G. Lucchetti, R. Lutzky Saute, H. Magdy Abd El Razek, F. G. Magnani, P. B. Mahajan, A. Majeed, A. Makki, R. Malekzadeh, A. A. Malik, N. Manafi, M. A. Mansournia, L. G. Mantovani, S. Martini, G. Mazzaglia, M. M. Mehndiratta, R. G. Menezes, A. Meretoja, A. G. Mersha, J. Miao Jonasson, B. Miazgowski, T. Miazgowski, I. M. Michalek, E. M. Mirrakhimov, Y. Mohammad, A. Mohamadian-Hafshejani, S. Mohammed, A. H. Mokdad, Y. Mokhayeri, M. Molokhia, M. A. Moni, A. A. Montasir, R. Moradzadeh, L. Morawska, J. Morze, W. Muruet, K. I. Musa, A. J. Nagarajan, M. Naghavi, S. Narasimha Swamy, B. R. Nascimento, R. I. Negoi, S. Neupane Kandel, T. H. Nguyen, B. Norrving, J. J. Noubiap, V. E. Nwatah, B. Oancea, O. O. Odukoya, A. T. Olagunju, H. Orru, M. O. Owolabi, J. R. Padubidri, A. Pana, T. Parekh, E.-C. Park, F. Pashazadeh Kan, M. Pathak, M. F. P. Peres, A. Perianayagam, T.-M. Pham, M. A. Piradov, V. Podder, S. Polinder, M. J. Postma, A. Pourshams, A. Radfar, A. Rafiei, A. Raggi, F. Rahim, V. Rahimi-Movaghar, M. Rahman, M. A. Rahman, A. M. Rahmani, N. Rajai, P. Ranasinghe, C. R. Rao, S. J. Rao, P. Rathi, D. L. Rawaf, S. Rawaf, M. B. Reitsma, V. Renjith, A. M. N. Renzaho, A. Rezapour, J. A. B. Rodriguez, L. Roever, M. Romoli, A. Rynkiewicz, S. Sacco, M. Sadeghi, S. Saeedi Moghaddam, A. Sahebkar, K. M. Saif-Ur-Rahman, R. Salah, M. Samaei, A. M. Samy, I. S. Santos, M. M. Santric-Milicevic, N. Sarrafzadegan, B. Sathian, D. Sattin, S. Schiavolin, M. P. Schlaich, M. I. A. Schmidt, A. E. Schutte, S. G. Sepanlou, A. Seylani, F. Sha, S. Shahabi, M. A. Shaikh, M. Shannawaz, M. S. R. Shawon, A. Sheikh, S. Sheikhbahaei, K. Shibuya, S. Siabani, D. A. S. Silva, J. A. Singh, J. K. Singh, V. Y. Skryabin, A. A. Skryabina, B. H. Sobaih, S. Stortecky, S. Stranges, E. G. Tadesse, I. U. Tarigan, M.-H. Temsah, Y. Teuschl, A. G. Thrift, M. Tonelli, M. R. Tovani-Palone, B. X. Tran, M. Tripathi, G. W. Tsegaye, A. Ullah, B. Unim, B. Unnikrishnan, A. Vakilian, S. Valadan Tahbaz, T. J. Vasankari, N. Venketasubramanian, D. Vervoort, B. Vo, V. Volovici, K. Vosoughi, G. T. Vu, L. G. Vu, H. A. Wafa, Y. Waheed, Y. Wang, T. Wijeratne, A. S. Winkler, C. D. A. Wolfe, M. Woodward, J. H. Wu, S. Wulf Hanson, X. Xu, L. Yadav, A. Yadollahpour, S. H. Yahyazadeh Jabbari, K. Yamagishi, H. Yatsuya, N. Yonemoto, C. Yu, I. Yunusa, M. S. Zaman, S. B. Zaman, M. Zamanian, R. Zand, A. Zandifar, M. S. Zastrozhin, A. Zastrozhina, Y. Zhang, Z.-J. Zhang, C. Zhong, Y. M. H. Zuniga and C. J. L. Murray, "Global, regional, and national burden of stroke and its risk factors, 1990-2019: A systematic analysis for the global burden of disease study 2019," *Lancet Neurol.* **20**(10), 795–820 (2021).
- [2] V. L. Feigin, M. Brainin, B. Norrving, S. Martins, R. L. Sacco, W. Hacke, M. Fisher, J. Pandian and P. Lindsay, "World stroke organization (WSO): Global stroke fact sheet 2022," *Int. J. Stroke* **17**(1), 18–29 (2022).
 - [3] U.S. Department of Veterans Affairs. "Management of Stroke Rehabilitation," 2024. <https://www.healthquality.va.gov/guidelines/rehab/stroke/index.asp> [accessed February 17, 2025].
 - [4] S. M. Hatem, G. Saussez, M. D. Faille, V. Prist, X. Zhang, D. Dispa and Y. Bleyenheuft, "Rehabilitation of motor function after stroke: A multiple systematic review focused on techniques to stimulate upper extremity recovery," *Front. Hum. Neurosci.* **10**, 442 (2016).
 - [5] Y. Bleyenheuft and A. M. Gordon, "Precision grip in congenital and acquired hemiparesis: Similarities in impairments and implications for neurorehabilitation," *Front. Hum. Neurosci.* **8**, 459 (2014).
 - [6] D. Corbetta, V. Sirtori, G. Castellini, L. Moja and R. Gatti, "Constraint-induced movement therapy for upper extremities in people with stroke," *Cochrane Database Syst. Rev.* **10**, CD004433 (2015).
 - [7] M. Etoom, M. Hawamdeh, Z. Hawamdeh, M. Alwardat, L. Giordani, S. Bacciu, C. Scarpini and C. Foti, "Constraint-induced movement therapy as a rehabilitation intervention for upper extremity in stroke patients: Systematic review and meta-analysis," *Int. J. Rehabil. Res.* **39**(3), 197–210 (2016).
 - [8] D. Pérez-Cruzado, J. A. Merchán-Baeza, M. González-Sánchez and A. I. Cuesta-Vargas, "Systematic review of mirror therapy compared with conventional rehabilitation in upper extremity function in stroke survivors," *Aust. Occup. Ther. J.* **64**, 91–112 (2017).
 - [9] A. Heller, D. T. Wade, V. A. Wood, A. Sunderland, R. L. Hower and E. Ward, "Arm function after stroke: Measurement and recovery over the first three months," *J. Neurol. Neurosurg. Psychiatry* **50**(6), 714–719 (1987).
 - [10] H. Nakayama, H. S. Jørgensen, H. O. Raaschou and T. S. Olsen, "Recovery of upper extremity function in stroke patients: The Copenhagen stroke study," *Arch. Phys. Med. Rehabil.* **75**(4), 394–398 (1994).
 - [11] G. Kwakkel, B. J. Kollen, J. van der Grond and A. J. H. Prevo, "Probability of regaining dexterity in the flaccid upper limb: Impact of severity of paresis and time since onset in acute stroke," *Stroke* **34**(9), 2181–2186 (2003).
 - [12] D. T. Wade, R. L. Hower, V. A. Wood, C. E. Skilbeck and H. M. Ismail, "The hemiplegic arm after stroke: Measurement and recovery," *J. Neurol. Neurosurg. Psychiatry* **46**(6), 521–524 (1983).
 - [13] A. Sunderland, D. Tinson, L. Bradley and R. L. Hower, "Arm function after stroke: An evaluation of grip strength as a measure of recovery and prognostic indicator," *J. Neurol. Neurosurg. Psychiatry* **52**(11), 1267–1272 (1989).
 - [14] R. Bertani, C. Melegari, M. C. De Cola, A. Bramanti, P. Bramanti and R. S. Calabrò, "Effects of robot-assisted upper limb rehabilitation in stroke patients: A systematic review with meta-analysis," *Neurol. Sci.* **38**(9), 1561–1569 (2017).
 - [15] A. C. Lo, P. D. Guarino, L. G. Richards, J. K. Haselkorn, G. F. Wittenberg, D. G. Federman, R. J. Ringer, T. H. Wagner, H. I. Krebs, B. T. Volpe, C. T. Bever Jr., D. M. Bravata, P. W. Duncan, B. H. Corn, A. D. Maffucci, S. E. Nadeau, S. S. Conroy, J. M. Powell, G. D. Huang and P. Peduzzi, "Robot-assisted therapy for long-term upper-limb impairment after stroke," *N. Engl. J. Med.* **362**(19), 1772–1783 (2010).
 - [16] S. Brochard, J. Robertson, B. Médée and O. Rémy-Néris, "What's new in new technologies for upper extremity rehabilitation?," *Curr. Opin. Neurol.* **23**(6), 683–687 (2010).

- [17] R. Gassert and V. Dietz, “Rehabilitation robots for the treatment of sensorimotor deficits: A neurophysiological perspective,” *J. Neuroeng. Rehabil.* **15**(1), 46 (2018).
- [18] R. C. Loureiro, W. S. Harwin, K. Nagai and M. Johnson, “Advances in upper limb stroke rehabilitation: A technology push,” *Med. Biol. Eng. Comput.* **49**(10), 1103–1118 (2011).
- [19] T. Otsuka, K. Kawaguchi, H. Kawamoto and Y. Sankai, “Development of upper-limb type HAL and reaching movement for meal-assistance,” *Proc. IEEE Int. Conf. Robot. Biomimetics*, **2011**, 883–888 (2011).
- [20] K. Kiguchi, M. H. Rahman, M. Sasaki and K. Teramoto, “Development of a 3-DOF mobile exoskeleton robot for human upper-limb motion assist,” *Robot. Auton. Syst.* **56**, 678–691 (2008).
- [21] G. Ivanova, S. Bulavintsev, J. H. Ryu and J. Poduraev, “Development of an exoskeleton system for elderly and disabled people,” *Int. Conf. Inf. Sci. Appl.* **2011**, 1–7 (2011).
- [22] M. A. Gull, M. Thoegersen, S. H. Bengtson, M. Mohammadi, L. N. S. A. Struijk, T. B. Moeslund, T. Bak, and S. Bai, “A 4-DOF upper limb exoskeleton for physical assistance: Design, modeling, control, and performance evaluation,” *Appl. Sci* **11**(13), 5865 (2021).
- [23] N. Hogan, H. I. Krebs, J. Charnnarong, P. Srikrishna and A. Sharon, “MIT-MANUS: A workstation for manual therapy and training. I,” *Proc. IEEE Int. Workshop Robot. Hum. Commun.* **1992**, 161–165 (1992).
- [24] F. Bovolenta, P. Sale, V. Dall’Armi, P. Clerici and M. Franceschini, “Robot-aided therapy for upper limbs in patients with stroke-related lesions,” *J. Neuroeng. Rehabil.* **8**, 18 (2011).
- [25] T. Hoesi, S. Iwagaki, J. Kai, K. Shigenobu and K. Kawahira, “Improvement of upper limb function for cervical myelopathy through repetitive facilitation exercise and reaching-robot therapy in an adult with cerebral palsy: A case study,” *Asian J. Occup. Ther.* **5**, 517–523 (2019).
- [26] D. Sui, J. Fan, H. Jin, X. Cai, J. Zhao and Y. Zhu, “Design of a wearable upper-limb exoskeleton for activities assistance of daily living,” *Proc. IEEE Int. Conf. Adv. Intell. Mechatronics (AIM)* **2017**, 1012–1017 (2017).
- [27] S. Christensen, S. Bai, S. Rafique, M. Isaksson, L. O’Sullivan, V. Power and G. S. Virk, “AXO-SUIT: A modular full-body exoskeleton for physical assistance,” *Mech. Des. Robot.* **2019**, 443–450 (2019).
- [28] Y. C. Liu, A. Botta, G. Quaglia and Y. Takeda, “Preliminary mechanical design of a wearable parallel-serial hybrid robot for wrist and forearm rehabilitation with consideration of joint misalignment compensation” *ROMANSY 24-Robot Design, Dynamics and Control*, **606**, 53–61 (2022).
- [29] Honda Motor Co., Ltd, “体重支持型歩行アシスト,” <https://global.honda.jp/robotics/weight/> [accessed January 23, 2025].
- [30] M. Bouzit, G. Burdea, G. Popescu and R. Boian, “The Rutgers Master II-New Design force-feedback glove,” *IEEE/ASME Trans. Mechatron.* **7**(2), 256–263 (2002).
- [31] R. Morita, M. Jiang, A. Botta, Y. Sugahara, G. Quaglia, M. Ceccarelli and Y. Takeda, “Design and preliminary testing of WELiBot: A wearable end-effector type upper limb assistive robot,” *Mech. Mach. Sci.* **164**, 67–75 (2024).
- [32] R. Morita, M. Jiang, A. Botta, Y. Sugahara, G. Quaglia, M. Ceccarelli and Y. Takeda, “Kinematic design and analysis of a wearable end-effector type upper limb assistive robot,” *Mech. Mach. Sci.* **148**, 561–570 (2023).
- [33] R. Suhaimi, K. S. Talha, K. Wan and M. A. Ariffin, “Design of movement sequences for arm rehabilitation of post-stroke,” *Proc. IEEE Int. Conf. Control Syst. Comput. Eng. (ICCSCE)* **2015**, 320–324 (2015).
- [34] Y. Hou and K. Kiguchi, “Generation of virtual tunnels for meal activity perception assist with an upper limb power assist exoskeleton robot,” *IEEE Access* **12**, 115137–115150 (2024).
- [35] L. Zhang, S. Guo and Q. Sun, “Development and assist-as-needed control of an end-effector upper limb rehabilitation robot,” *Appl. Sci.* **10**(19), 6684 (2020).
- [36] M. Higuchi and T. Ogasawara, “Development of a human symbiotic assist arm PAS-Arm,” *Proc. IEEE Int. Conf. Rehabil. Robot. (ICORR)* **2009**, 491–497 (2009).
- [37] J. L. Pons, R. Ceres, E. Rocon, S. Levin, I. Markovitz, B. Saro, D. Reynaerts, W. Van Moorlegghem and L. Bueno, “Virtual reality training and EMG control of the MANUS hand prosthesis,” *Robotica* **23**(3), 311–317 (2025).
- [38] J. Zhang, H. Zeng, X. Li, G. Xu, Y. Li and A. Song, “Bayesian optimization for assist-as-needed controller in robot-assisted upper limb training based on energy information,” *Robotica* **41**(10), 3101–3115 (2023).

Cite this article: M. Jiang, R. Morita, Y. Wang, T. Shinoda, Q. Meng, A. Botta, G. Quaglia, M. Ceccarelli and Y. Takeda (2025). “Design and analysis of WELiBot for assisting upper limb motion”, *Robotica* **43**, 3921–3941. <https://doi.org/10.1017/S0263574725102476>