

NDIR Optical E-Nose for Contactless Diabetes Monitoring and Diagnosis: An integrated HW-AI Framework

Original

NDIR Optical E-Nose for Contactless Diabetes Monitoring and Diagnosis: An integrated HW-AI Framework / Giarrusso, G.S., Olmo, G., Gumiero, A.. - ELETTRONICO. - (2026), pp. 79-80. (2026 IEEE International Conference on Pervasive Computing and Communications Workshops and other Affiliated Events (PerCom Workshops) Pisa (Italy) 16-20 March 2026) [10.1109/PerComWorkshops68308.2026.11585237].

Availability:

This version is available at: 11583/3013108 since: 2026-07-16T11:20:53Z

Publisher:

IEEE

Published

DOI:10.1109/PerComWorkshops68308.2026.11585237

Terms of use:

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

IEEE postprint/Author's Accepted Manuscript

©2026 IEEE. Personal use of this material is permitted. Permission from IEEE must be obtained for all other uses, in any current or future media, including reprinting/republishing this material for advertising or promotional purposes, creating new collecting works, for resale or lists, or reuse of any copyrighted component of this work in other works.

(Article begins on next page)

NDIR Optical E-Nose for Contactless Diabetes Monitoring and Diagnosis: An integrated HW-AI Framework

Gabriele Salvatore Giarrusso¹

Supervised by: Gabriella Olmo¹ and Alessandro Gumiero²

¹Department of Control and Computer Engineering at Politecnico di Torino

²STMicroelectronics

gabriele.giarrusso@polito.it

Abstract—Non-invasive and widely accessible methods for Diabetes Mellitus (DM) screening and monitoring are crucial to address the limitations of conventional blood-based assays. Exhaled breath (EB) analysis provides a contactless, real-time alternative by detecting disease-associated changes in Volatile Organic Compounds (VOCs). This PhD research work introduces a compact Non-dispersive Infrared (NDIR)-based optical Electronic Nose (E-nose) for rapid VOCs detection. Preliminary laboratory tests using a general-purpose IR sensor and a MEMS IR emitter showed feasibility for detecting low VOCs concentrations. Expanding the system with nonspecific near-/mid-IR, MOS, and ambient sensors could further enhance performance, enabling AI-driven multi-modal strategies for DM-oriented breath analysis.

Index Terms—Diabetes monitoring, Contactless sensing, AI platform, IR Breath Analysis, Biomedical Sensors

I. INTRODUCTION

Diabetes Mellitus (DM), a chronic disorder of insulin production/action, affects hundreds of millions globally (nearly half undiagnosed), with early detection and monitoring critical to prevent severe complications [1]. Traditional blood-based diagnostics, such as finger-prick glucose and HbA1c tests, are invasive and inconvenient, motivating non-invasive approaches [1]. Exhaled Breath (EB) analysis enables contactless monitoring by detecting Volatile Organic Compounds (VOCs) linked to metabolic processes, including acetone (main DM biomarker) isoprene, ethanol, isopropanol, ammonia, and sulfur-containing compounds, which can be analyzed in clusters to improve diagnostic accuracy and reduce individual variability [2], [3], [4], [5]. Detection methods range from highly sensitive but bulky mass-spectrometry techniques, namely Gas Chromatography (GC-MS), Proton Transfer Reaction (PTR-MS), Selected Ion Flow Tube (SIFT-MS), to portable chemical sensors (MOS, electrochemical), which are rapid but limited in selectivity and affected by cross-sensitivity [2]. Among Infrared (IR)-based methods, Non-Dispersive Infrared (NDIR) spectroscopy provides the best balance between cost, specificity, and portability, though challenges remain regarding water interference and sensitivity [2].

This research work, conducted as part of a PhD project, presents a compact NDIR-based optical Electronic Nose (E-nose) for rapid, non-invasive detection of DM-related VOCs in EB (Figure 1), suggesting promising results. Further integration of near-/mid-IR, MOS, and ambient sensor data would support an AI-based multi-modal platform that leverages both spectroscopic specificity and E-nose versatility,




Fig. 1. Framework pipeline for non-invasive DM assessment

enabling data-driven approaches for enhanced, contactless, real-time DM monitoring and diagnosis. As a proof-of-concept, this work demonstrates preliminary system behavior, assessing feasibility for detecting low VOCs concentrations under controlled laboratory conditions.

II. MATERIALS & METHODS

The developed system employs a MEMS IR source (EMIRS200, Axetris) operating over 2–14 μ m and driven at 50Hz with a 62% duty cycle, together with a general-purpose Thermal MOS (TMOS) sensor (STHS34PF80, STMicroelectronics) covering 5–20 μ m and originally intended for presence detection, providing an effective 1Hz output through the averaging of 128 samples per measurement [6]. Both components are integrated in a modular, 3D-printed Polyamide12 (PA12) chamber featuring vapor-smoothed walls to increase reflectivity and reduce material porosity. The chamber (23.5mm internal diameter, 10cm path length, 43.37mL volume, with $\tau \approx 6.5$ s when flow=400mL/min) is housed in aluminum-coated boxes to reduce both environmental and electrical noise. Controlled




Fig. 2. Multi-modal platform for VOCs detection

TABLE I
MFC SETTINGS FOR CONTROLLED VOCs INJECTION

VOCs	Gas [ppm]	Base	C [ppm]	Flow [SCCM]	T G/A [s]
C ₃ H ₆ O	1000	O ₂ , N ₂	500, 100, 10	500	41/62
NH ₃	1000	N ₂	400, 200, 100	400	200/800
H ₂ S	500	N ₂	225, 100, 50	400	500/1000

VOCs - i.e. Acetone (C₃H₆O), Ammonia (NH₃), Hydrogen Sulfide (H₂S) - were injected via Mass Flow Controller (MFC) (Table I) at 1.5bar, alternating with clean air, without temperature or humidity control (T_{amb} 21–23°C, RH_{amb} 20–50%). TMOS signals were smoothed (moving average, $M=50$) and segmented by response and recovery windows.

III. RESULTS

Window-related data are used to compute the relative response

$$S\% = \frac{R_O - R_G}{R_O} \times 100$$

which compares gas-exposure output (R_G) to baseline (R_O), quantifying absorption at the target wavelength and providing a standardized metric of sensitivity, linearity, and reproducibility across gases (Figure 3). H₂S produced the strongest response (2.31% at 225ppm), exceeding C₃H₆O (1.66% at 500ppm) and NH₃ (0.67% at 400ppm); at equal concentrations (100ppm), the trend persisted (0.81%, 0.55%, 0.30% respectively). These results indicate that the system can detect C₃H₆O and NH₃ down to 100ppm and H₂S to 50ppm, with preliminary concentration-response trends showing near-linear correlations ($r=0.99 \pm 0.10$, $R^2=0.98 \pm 0.03$) (Figure 3), suggesting the feasibility of the proposed system for low VOCs traces detection.

IV. DISCUSSION & FUTURE PERSPECTIVES

DM is a widespread chronic disorder causing hyperglycemia and complications, with early assessment essential for management; however, current blood-based tests are invasive and unsuitable. EB analysis offers a non-invasive alternative, but existing methods are limited by cost, size, or cross-sensitivity;




Fig. 3. Preliminary results in lab-settings for C₃H₆O (Blue), H₂S (Pink), and NH₃ (Green) respectively

In this context, NDIR spectroscopy provides the best compromise.

This PhD research work suggests the feasibility of a compact, non-invasive NDIR-based optical E-nose for detecting DM-related VOCs in EB. Preliminary lab-experiments using a general-purpose IR sensor and a MEMS IR emitter showed reliable detection of reduced C₃H₆O, NH₃, and H₂S concentrations, showing near-linear concentration–response behavior and clear relative responses, even under early, non-optimized experimental conditions.

Limitations of the current system include inconsistencies in driving modes, chamber conditioning, and MFC timing, which reduce SNR, compromise reproducibility, and hinder comparative analysis. Future work should refine the experimental conditions, increase concentration granularity, standardize MFC protocols, employ inert chamber materials, and further characterize emitter-sensor behavior. The platform would further benefit from integrating wavelength-selective detectors (targeting C₃H₆O absorption peaks at 5.71μm and 8.2μm), broadband reference TMOS sensors (2–20μm), MOS-based chemical sensors, and auxiliary environmental monitoring (pressure, temperature, relative humidity) to form a multi-modal sensing architecture. Such enhancements would enable more robust signal extraction by isolating VOCs-related variations, mitigating water-vapor and temperature interference, and incorporating broadband IR information to improve sensitivity and specificity.

Leveraging software-driven analysis rather than relying solely on hardware specificity, these multi-modal data streams can be processed through machine-learning algorithms for both classification and regression, ultimately supporting the development of a practical, contactless, and rapid point-of-care VOCs detection system for DM monitoring and diagnosis.

ACKNOWLEDGMENT

The drafting and publication of this work took place in the context of the project PNRR-NGEU which has received funding from the MUR-DM 117/2023 (Italy).

REFERENCES

- [1] V. Saasa, T. Malwela, M. Beukes, M. Mokgotho, C.-P. Liu, and B. Mwakikunga, “Sensing technologies for detection of acetone in human breath for diabetes diagnosis and monitoring,” *Diagnostics*, vol. 8, no. 1, 2018.
- [2] U. B. Ranu, M. A. Rahman, S. Sriram, and P. B. Agarwal, “Infrared non-invasive exhaled biomarker sensing: A review,” *Advanced Sensor Research*, vol. 3, no. 1, p. 2300085, 2024.
- [3] K. Dixit, S. Fardindoost, A. Ravishankara, N. Tasnim, and M. Hoorfar, “Exhaled breath analysis for diabetes diagnosis and monitoring: Relevance, challenges and possibilities,” *Biosensors*, vol. 11, no. 12, 2021.
- [4] Y. V. Kistenev, A. V. Borisov, V. S. Zasedatel, and L. V. Spirina, “Diabetes noninvasive diagnostics and monitoring through volatile biomarkers analysis in the exhaled breath using optical absorption spectroscopy,” *Journal of Biophotonics*, vol. 16, no. 12, p. e202300198, 2023.
- [5] A. Sharma, R. Kumar, and P. Varadwaj, “Smelling the disease: Diagnostic potential of breath analysis,” *Mol. Diagn. Ther.*, vol. 27, pp. 321–347, May 2023.
- [6] E. Moisello, M. E. Castagna, A. L. Malfa, G. Bruno, P. Malcovati, and E. Bonizzoni, “Reference temperature sensor for tmoss-based thermal detectors,” *IEEE Access*, vol. 11, pp. 96594–96602, 2023.